

LB115003


245.1

INSTRUMENT ID: LV1

Linear

 μ Abs.:

28429



A= 0.0000e+000

B= 3.4948e-004

C= -9.0315e-004

Rho= 0.9999193

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/01/12
0.0	0.000	-0.004	-0.004	-10	0.000	-10					-
0.05	0.050	0.053	0.003	154	0.0 %	154	0				6
0.2	0.200	0.198	-0.002	569	0.0 %	569					-1
2.5	2.500	2.443	-0.057	6993	0.0 %	6993					-2
5.0	5.000	5.065	0.065	14496	0.0 %	14496					1
7.5	7.500	7.561	0.061	21638	0.0 %	21638					1
10.0	10.000	9.934	-0.066	28429	0.0 %	28429					-1

LB115003 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Jun 2021 12:03:40

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.0000	245.1	PPB	17 Jun 2021 12:08:48	S	Std
0.05 - 1	50.05	154	-	0.0500	245.1	PPB	17 Jun 2021 12:11:05	S	Std
0.2 - 1	50.2	569	-	0.2000	245.1	PPB	17 Jun 2021 12:13:22	S	Std
2.5 - 1	52.5	6993	-	2.5000	245.1	PPB	17 Jun 2021 12:15:39	S	Std
5.0 - 1	55	14496	-	5.0000	245.1	PPB	17 Jun 2021 12:17:55	S	Std
7.5 - 1	57.5	21638	-	7.5000	245.1	PPB	17 Jun 2021 12:20:11	S	Std
10.0 - 1	510	28429	-	10.0000	245.1	PPB	17 Jun 2021 12:34:46	S	Std
ICV66 - 1	ICV66	11073	3.8689	-245.1		PPB	17 Jun 2021 12:37:29	U	SMPL
ICB66 - 1	ICB66	-120	-0.0428	-245.1		PPB	17 Jun 2021 12:39:44	U	SMPL
CCV19 - 1	CCV19	13931	4.8677	-245.1		PPB	17 Jun 2021 12:41:59	U	SMPL
CCB19 - 1	CCB19	-113	-0.0404	-245.1		PPB	17 Jun 2021 12:44:14	U	SMPL
CRA - 1	CRA	150	0.0515	-245.1		PPB	17 Jun 2021 12:52:05	U	SMPL
HighStd - 1	HighStd	28437	9.9372	-245.1		PPB	17 Jun 2021 12:56:59	U	SMPL
ChkStd - 1	ChkStd	19676	6.8754	-245.1		PPB	17 Jun 2021 13:03:49	U	SMPL
PB137177BL - 1	PBW	-122	-0.0435	-245.1		PPB	17 Jun 2021 13:06:05	U	SMPL
PB137177BS - 1	LCSW	10564	3.6910	-245.1		PPB	17 Jun 2021 13:10:54	U	SMPL
M2262-10 - 1	LOD-MDL-WATER-04-QT2-2021	457	0.1588	-245.1		PPB	17 Jun 2021 13:15:44	U	SMPL
M2262-12 - 1	MDL-WATER-06-QT2-2021	143	0.0491	-245.1		PPB	17 Jun 2021 13:18:00	U	SMPL
CCV20 - 1	CCV20	14024	4.9002	-245.1		PPB	17 Jun 2021 13:22:49	U	SMPL
CCB20 - 1	CCB20	-105	-0.0376	-245.1		PPB	17 Jun 2021 13:25:05	U	SMPL