

LB 115006


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38750



A= 0.0000e+000

B= 2.5776e-004 slope

C= 1.5608e-002 y -intercept

Rho= 0.9999615

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	900
0.00	0.000	0.025	0.025	38	0.000	38					10
0.05	0.050					207					-2
0.20	0.200	0.219	0.019	789	0.0 %	789					-1
2.50	2.500	2.461	-0.039	9488	0.0 %	9488					1
5.00	5.000	4.952	-0.048	19152	0.0 %	19152					0
7.50	7.500	7.539	0.039	29186	0.0 %	29186					
10.0	10.000	10.004	0.004	38750	0.0 %	38750					

LB115006
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	38	-		0 SFAM01.1	PPB	6/17/2021 14:17	Std
	0.2 S01	789	-		0.2 SFAM01.1	PPB	6/17/2021 14:20	Std
	2.5 S02	9488	-		2.5 SFAM01.1	PPB	6/17/2021 14:22	Std
	5 S03	19152	-		5 SFAM01.1	PPB	6/17/2021 14:24	Std
	7.5 S04	29186	-		7.5 SFAM01.1	PPB	6/17/2021 14:27	Std
	10 S05	38750	-		10 SFAM01.1	PPB	6/17/2021 14:30	Std
ICV046	ICV046	14467	3.7446 -		SFAM01.1	PPB	6/17/2021 14:33	SMPL
ICB046	ICB046	-70	-0.0024 -		SFAM01.1	PPB	6/17/2021 14:35	SMPL
CCV021	CCV021	19409	5.0184 -		SFAM01.1	PPB	6/17/2021 14:38	SMPL
CCB021	CCB021	-93	-0.0084 -		SFAM01.1	PPB	6/17/2021 14:40	SMPL
PB137176BL	PBW176	22	0.0213 -		SFAM01.1	PPB	6/17/2021 14:42	SMPL
M2694-01	BGHJ0	47	0.0277 -		SFAM01.1	PPB	6/17/2021 14:44	SMPL
M2694-02	BGHJ2	615	0.1741 -		SFAM01.1	PPB	6/17/2021 14:47	SMPL
M2694-03	BGHJ3	38	0.0254 -		SFAM01.1	PPB	6/17/2021 14:49	SMPL
M2694-04	BGHH8	48	0.028 -		SFAM01.1	PPB	6/17/2021 14:51	SMPL
M2694-05	BGHH8D	43	0.0267 -		SFAM01.1	PPB	6/17/2021 14:53	SMPL
M2694-06	BGHH8S	3801	0.9953 -		SFAM01.1	PPB	6/17/2021 14:56	SMPL
M2694-07	BGHH9	40	0.0259 -		SFAM01.1	PPB	6/17/2021 14:58	SMPL
M2694-08	BGHJ1	23	0.0215 -		SFAM01.1	PPB	6/17/2021 15:00	SMPL
CCV022	CCV022	19028	4.9202 -		SFAM01.1	PPB	6/17/2021 15:03	SMPL
CCB022	CCB022	-103	-0.0109 -		SFAM01.1	PPB	6/17/2021 15:05	SMPL