


LB 115008

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0017e-004

C= 2.1443e-002

Rho= 0.9999665

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	0.032	0.032	54	0.000	54					1
0.2	0.200	0.214	0.014	962	0.0 %	962					7
2.5	2.500	2.450	-0.050	12131	0.0 %	12131					-2
5.0	5.000	4.997	-0.003	24859	0.0 %	24859					0
7.5	7.500	7.474	-0.026	37233	0.0 %	37233					0
10.0	10.000	10.032	0.032	50012	0.0 %	50012					0

LB115008 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jun 2021 15:24:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	54	-	0.0000	7471B	PPB	17 Jun 2021 15:26:28	S	Std
0.2 - 1	SD	962	-	0.2000	7471B	PPB	17 Jun 2021 15:28:45	S	Std
2.5 - 1	SD	12131	-	2.5000	7471B	PPB	17 Jun 2021 15:31:02	S	Std
5.0 - 1	SD	24859	-	5.0000	7471B	PPB	17 Jun 2021 15:33:19	S	Std
7.5 - 1	SD	37233	-	7.5000	7471B	PPB	17 Jun 2021 15:35:34	S	Std
10.0 - 1	SD	50012	-	10.0000	7471B	PPB	17 Jun 2021 15:37:51	S	Std
ICV67 - 1	ICV67	19231	3.8709	-7471B	PPB	PPB	17 Jun 2021 15:40:34	U	SMPL
ICB67 - 1	ICB67	-102	0.0010	-7471B	PPB	PPB	17 Jun 2021 15:42:49	U	SMPL
CCV23 - 1	CCV23	24425	4.9106	-7471B	PPB	PPB	17 Jun 2021 15:45:05	U	SMPL
CCB23 - 1	CCB23	-91	0.0032	-7471B	PPB	PPB	17 Jun 2021 15:47:20	U	SMPL
CRA - 1	CRA	915	0.2046	-7471B	PPB	PPB	17 Jun 2021 15:49:35	U	SMPL
HighStd - 1	HighStd	48556	9.7408	-7471B	PPB	PPB	17 Jun 2021 15:51:50	U	SMPL
ChkStd - 1	ChkStd	34345	6.8962	-7471B	PPB	PPB	17 Jun 2021 15:54:06	U	SMPL
PB137172BL - 1	PBS	-192	-0.0170	-7471B	PPB	PPB	17 Jun 2021 15:56:24	U	SMPL
PB137172BS - 1	LCSS	16697	3.3637	-7471B	PPB	PPB	17 Jun 2021 15:58:41	U	SMPL
M2719-01 - 1	OK-03-061521	1170	0.2556	-7471B	PPB	PPB	17 Jun 2021 16:00:57	U	SMPL
M2734-01 - 1	CWSP	1362	0.2941	-7471B	PPB	PPB	17 Jun 2021 16:03:13	U	SMPL
M2735-01 - 1	124027	3075	0.6370	-7471B	PPB	PPB	17 Jun 2021 16:05:29	U	SMPL
M2735-03 - 1	124035	9894	2.0019	-7471B	PPB	PPB	17 Jun 2021 16:07:47	U	SMPL
M2735-05 - 1	124055	3805	0.7831	-7471B	PPB	PPB	17 Jun 2021 16:10:02	U	SMPL
CCV24 - 1	CCV24	23235	4.6724	-7471B	PPB	PPB	17 Jun 2021 16:12:17	U	SMPL
CCB24 - 1	CCB24	-101	0.0012	-7471B	PPB	PPB	17 Jun 2021 16:14:32	U	SMPL
M2736-01 - 1	PL-01-061621	2173	0.4564	-7471B	PPB	PPB	17 Jun 2021 16:16:48	U	SMPL
M2737-01 - 1	SU-03-061621	1311	0.2839	-7471B	PPB	PPB	17 Jun 2021 16:19:04	U	SMPL
M2738-01 - 1	HD-01-061621	468	0.1151	-7471B	PPB	PPB	17 Jun 2021 16:21:19	U	SMPL
M2739-01 - 1	VNJ-214	1741	0.3699	-7471B	PPB	PPB	17 Jun 2021 16:23:35	U	SMPL
M2746-02 - 1	NEW-DORP-2	2616	0.5451	-7471B	PPB	PPB	17 Jun 2021 16:25:50	U	SMPL
M2748-01 - 1	SB-3A	729	0.1674	-7471B	PPB	PPB	17 Jun 2021 16:28:06	U	SMPL
M2748-02 - 1	SB-3B	578	0.1371	-7471B	PPB	PPB	17 Jun 2021 16:30:22	U	SMPL
M2748-05 - 1	SB-5A	3612	0.7445	-7471B	PPB	PPB	17 Jun 2021 16:32:40	U	SMPL
M2748-07 - 1	SB-8A	6886	1.3998	-7471B	PPB	PPB	17 Jun 2021 16:34:57	U	SMPL
M2748-09 - 1	SB-8-AOC-2	526	0.1267	-7471B	PPB	PPB	17 Jun 2021 16:37:13	U	SMPL
CCV25 - 1	CCV25	24613	4.9482	-7471B	PPB	PPB	17 Jun 2021 16:52:17	U	SMPL
CCB25 - 1	CCB25	-116	-0.0018	-7471B	PPB	PPB	17 Jun 2021 16:54:32	U	SMPL
M2748-10 - 1	SB-9A	3165	0.6550	-7471B	PPB	PPB	17 Jun 2021 16:56:48	U	SMPL
M2748-12 - 1	SB-10A	1206	0.2628	-7471B	PPB	PPB	17 Jun 2021 16:59:03	U	SMPL
M2748-14 - 1	SB-10-AOC-2	511	0.1237	-7471B	PPB	PPB	17 Jun 2021 17:01:19	U	SMPL
M2748-15 - 1	SB-6A	5685	1.1594	-7471B	PPB	PPB	17 Jun 2021 17:03:34	U	SMPL
M2748-15DUP - 1	SB-6ADUP	6627	1.3480	-7471B	PPB	PPB	17 Jun 2021 17:05:50	U	SMPL
M2748-15MS - 1	SB-6AMS	26293	5.2845	-7471B	PPB	PPB	17 Jun 2021 17:08:06	U	SMPL
M2748-15MSD - 1	SB-6AMSD	27266	5.4792	-7471B	PPB	PPB	17 Jun 2021 17:10:23	U	SMPL
M2748-15LX5 - 1		1038	0.2292	-7471B	PPB	PPB	17 Jun 2021 17:12:40	U	SMPL
M2748-15A - 1		26855	5.3970	-7471B	PPB	PPB	17 Jun 2021 17:14:58	U	SMPL
CCV26 - 1	CCV26	23320	4.6894	-7471B	PPB	PPB	17 Jun 2021 17:17:14	U	SMPL
CCB26 - 1	CCB26	-65	0.0084	-7471B	PPB	PPB	17 Jun 2021 17:19:29	U	SMPL