



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115009

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/17/2021	13:20
2	CAL2	1	Water	NA	NA	20.2	7.00	06/17/2021	13:21
3	CAL3	1	Water	NA	NA	20.3	10.02	06/17/2021	13:22
4	ICV	1	Water	NA	NA	20.4	7.01	06/17/2021	13:24
5	CCV1	1	Water	NA	NA	20.3	2.01	06/17/2021	13:25
6	M2764-02	1	Solid	20.02	20	21.3	6.88	06/17/2021	13:26
7	M2764-02DUP	1	Solid	20.03	20	21.4	6.89	06/17/2021	13:27
8	M2764-04	1	Solid	20.05	20	22.1	7.64	06/17/2021	13:30
9	CCV2	1	Water	NA	NA	20.4	12.02	06/17/2021	13:31

VB 115009

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2764 WorkList ID : 149928 Department : Wet-Chemistry Date : 06-17-2021 13:05:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/16/2021	Solid	M2764-02	Corrosivity	Cool 4 deg C	NOBI03	F51	PHC-250EM-001	06/16/2021	9045D
06/16/2021	Solid	M2764-04	Corrosivity	Cool 4 deg C	NOBI03	F51	PHC-250EM-002	06/16/2021	9045D

Date/Time 06-17-21 13:10
 Raw Sample Received by: JO (wet Chem)
 Raw Sample Relinquished by: JT (sm)

Date/Time 06-17-21 16:35
 Raw Sample Received by: JT (sm)
 Raw Sample Relinquished by: JO (wet Chem)