

LB115015


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39672



A= 0.0000e+000

B= 2.5271e-004

C= 8.2490e-003

Rho= 0.9998941

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.013	0.013	17	0.000	17					12
0.2	0.200	0.224	0.024	852	0.0 %	852					-3
2.5	2.500	2.431	-0.069	9586	0.0 %	9586					1
5.0	5.000	5.073	0.073	20041	0.0 %	20041					-1
7.5	7.500	7.427	-0.073	29356	0.0 %	29356					0
10.0	10.000	10.034	0.034	39672	0.0 %	39672					

LB115015 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 18 Jun 2021 11:16:47							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	17	-	0.0007470A	PPB		18 Jun 2021 11:19:37	S	Std
0.2 - 1	50.2	852	-	0.20007470A	PPB		18 Jun 2021 11:21:53	S	Std
2.5 - 1	52.5	9586	-	2.50007470A	PPB		18 Jun 2021 11:24:10	S	Std
5.0 - 1	55	20041	-	5.00007470A	PPB		18 Jun 2021 11:26:26	S	Std
7.5 - 1	57.5	29356	-	7.50007470A	PPB		18 Jun 2021 11:28:43	S	Std
10.0 - 1	57.8	39672	-	10.00007470A	PPB		18 Jun 2021 11:30:59	S	Std
ICV68 - 1	ICV68	15348	3.8868	-7470A	PPB		18 Jun 2021 11:33:40	U	SMPL
ICB68 - 1	ICB68	-145	-0.0284	-7470A	PPB		18 Jun 2021 11:35:55	U	SMPL
CCV27 - 1	CCV27	19763	5.0025	-7470A	PPB		18 Jun 2021 11:38:10	U	SMPL
CCB27 - 1	CCB27	-144	-0.0281	-7470A	PPB		18 Jun 2021 11:40:25	U	SMPL
CRA - 1	CRA	733	0.1935	-7470A	PPB		18 Jun 2021 11:42:40	U	SMPL
HighStd - 1	HighStd	36105	9.1323	-7470A	PPB		18 Jun 2021 11:44:54	U	SMPL
ChkStd - 1	ChkStd	25291	6.3995	-7470A	PPB		18 Jun 2021 11:47:10	U	SMPL
PB137174BL - 1	PBW	-208	-0.0443	-7470A	PPB		18 Jun 2021 11:49:27	U	SMPL
PB137174BS - 1	LCSW	15137	3.8335	-7470A	PPB		18 Jun 2021 11:54:17	U	SMPL
M2710-02 - 1	WC-1	-128	-0.0241	-7470A	PPB		18 Jun 2021 11:56:32	U	SMPL
M2715-01 - 1	33248	-6	0.0067	-7470A	PPB		18 Jun 2021 11:58:49	U	SMPL
M2735-02 - 1	124027	0	0.0082	-7470A	PPB		18 Jun 2021 12:01:05	U	SMPL
M2735-04 - 1	124035	27	0.0151	-7470A	PPB		18 Jun 2021 12:03:22	U	SMPL
M2735-06 - 1	124055	1	0.0085	-7470A	PPB		18 Jun 2021 12:05:37	U	SMPL
CCV28 - 1	CCV28	19972	5.0553	-7470A	PPB		18 Jun 2021 12:07:51	U	SMPL
CCB28 - 1	CCB28	-130	-0.0246	-7470A	PPB		18 Jun 2021 12:10:07	U	SMPL
M2746-03 - 1	NEW-DORP-2	-25	0.0019	-7470A	PPB		18 Jun 2021 12:12:24	U	SMPL
M2746-03DUP - 1	NEW-DORP-2DUP	-12	0.0052	-7470A	PPB		18 Jun 2021 12:14:39	U	SMPL
M2746-03MS - 1	NEW-DORP-2MS	15816	4.0051	-7470A	PPB		18 Jun 2021 12:19:26	U	SMPL
M2746-03MSD - 1	NEW-DORP-2MSD	15839	4.0109	-7470A	PPB		18 Jun 2021 12:21:42	U	SMPL
PB137175BL - 1	PBW	-121	-0.0223	-7470A	PPB		18 Jun 2021 12:23:57	U	SMPL
PB137175BS - 1	LCSW	15490	3.9227	-7470A	PPB		18 Jun 2021 12:26:13	U	SMPL
M2740-01 - 1	0342-FIELD-BLANK	-129	-0.0244	-7470A	PPB		18 Jun 2021 12:28:29	U	SMPL
M2741-02 - 1	0336-SITE-WATER	55	0.0221	-7470A	PPB		18 Jun 2021 12:30:46	U	SMPL
M2741-02DUP - 1	0336-SITE-WATERDUP	60	0.0234	-7470A	PPB		18 Jun 2021 12:33:03	U	SMPL
M2741-02MS - 1	0336-SITE-WATERMS	15574	3.9439	-7470A	PPB		18 Jun 2021 12:35:18	U	SMPL
CCV29 - 1	CCV29	19795	5.0106	-7470A	PPB		18 Jun 2021 12:37:34	U	SMPL
CCB29 - 1	CCB29	-143	-0.0279	-7470A	PPB		18 Jun 2021 12:39:49	U	SMPL
M2741-02MSD - 1	0336-SITE-WATERMSD	14866	3.7650	-7470A	PPB		18 Jun 2021 12:42:05	U	SMPL
PB137190BL - 1	PBW	-100	-0.0170	-7470A	PPB		18 Jun 2021 12:44:20	U	SMPL
PB137190BS - 1	LCSW	16311	4.1302	-7470A	PPB		18 Jun 2021 12:49:13	U	SMPL
M2762-02 - 1	WC_AQUEOUS_DBD	175	0.0525	-7470A	PPB		18 Jun 2021 12:51:28	U	SMPL
M2762-02DUP - 1	WC_AQUEOUS_DBDDUP	213	0.0621	-7470A	PPB		18 Jun 2021 12:53:44	U	SMPL
M2762-02MS - 1	WC_AQUEOUS_DBDMS	15272	3.8676	-7470A	PPB		18 Jun 2021 12:58:33	U	SMPL
M2762-02MSD - 1	WC_AQUEOUS_DBDMSD	15484	3.9212	-7470A	PPB		18 Jun 2021 13:00:49	U	SMPL
CCV30 - 1	CCV30	19881	5.0323	-7470A	PPB		18 Jun 2021 13:03:06	U	SMPL
CCB30 - 1	CCB30	-86	-0.0135	-7470A	PPB		18 Jun 2021 13:05:23	U	SMPL
PB137174TB - 1	PB137150TB	-19	0.0034	-7470A	PPB		18 Jun 2021 13:07:39	U	SMPL
PB137190TB - 1	PB137178TB	68	0.0254	-7470A	PPB		18 Jun 2021 13:09:54	U	SMPL
M2746-03LX5 - 1		64	0.0244	-7470A	PPB		18 Jun 2021 13:12:09	U	SMPL
M2746-03A - 1		15718	3.9803	-7470A	PPB		18 Jun 2021 13:14:25	U	SMPL
M2741-02LX5 - 1		-40	-0.0019	-7470A	PPB		18 Jun 2021 13:16:40	U	SMPL
M2741-02A - 1		15696	3.9748	-7470A	PPB		18 Jun 2021 13:18:56	U	SMPL
M2762-02LX5 - 1		70	0.0259	-7470A	PPB		18 Jun 2021 13:21:11	U	SMPL
M2762-02A - 1		17075	4.3232	-7470A	PPB		18 Jun 2021 13:25:32	U	SMPL
CCV31 - 1	CCV31	20440	5.1736	-7470A	PPB		18 Jun 2021 13:27:48	U	SMPL
CCB31 - 1	CCB31	-142	-0.0276	-7470A	PPB		18 Jun 2021 13:30:05	U	SMPL