


LB115021

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0951e-004 *slope*

C= 7.6111e-003 *+intercept*

Rho= 0.9999751

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/ol)</i>
0.00	0.000	0.012	0.012	21	0.000	21					<i>2</i>
0.05	0.050					207					<i>2</i>
0.20	0.200	0.203	0.003	932	0.0 %	932					<i>-2</i>
2.50	2.500	2.456	-0.044	11686	0.0 %	11686					<i>1</i>
5.00	5.000	5.043	0.043	24033	0.0 %	24033					<i>0</i>
7.50	7.500	7.489	-0.011	35706	0.0 %	35706					<i>0</i>
10.0	10.000	9.998	-0.002	47682	0.0 %	47682					<i>0</i>

LB115021
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	21	-		0 SFAM01.1	PPB	6/18/2021 13:44	Std
	0.2 S01	932	-		0.2 SFAM01.1	PPB	6/18/2021 13:47	Std
	2.5 S02	11686	-		2.5 SFAM01.1	PPB	6/18/2021 13:49	Std
	5 S03	24033	-		5 SFAM01.1	PPB	6/18/2021 13:51	Std
	7.5 S04	35706	-		7.5 SFAM01.1	PPB	6/18/2021 13:54	Std
	10 S05	47682	-		10 SFAM01.1	PPB	6/18/2021 13:56	Std
ICV047	ICV047	18692	3.9239 -		SFAM01.1	PPB	6/18/2021 13:59	SMPL
ICB047	ICB047	-96	-0.0125 -		SFAM01.1	PPB	6/18/2021 14:01	SMPL
CCV032	CCV032	24187	5.0751 -		SFAM01.1	PPB	6/18/2021 14:03	SMPL
CCB032	CCB032	-99	-0.0131 -		SFAM01.1	PPB	6/18/2021 14:05	SMPL
PB137204BL	PBS204	22	0.0122 -		SFAM01.1	PPB	6/18/2021 14:08	SMPL
M2696-01	BGHJ8	3037	0.6439 -		SFAM01.1	PPB	6/18/2021 14:10	SMPL
M2696-02	BGHK0	1385	0.2978 -		SFAM01.1	PPB	6/18/2021 14:12	SMPL
M2696-03	BGHK5	2206	0.4698 -		SFAM01.1	PPB	6/18/2021 14:14	SMPL
M2696-04	BGHK6	433	0.0983 -		SFAM01.1	PPB	6/18/2021 14:17	SMPL
M2696-05	BGHJ6	1760	0.3764 -		SFAM01.1	PPB	6/18/2021 14:19	SMPL
M2696-06	BGHJ6D	1720	0.368 -		SFAM01.1	PPB	6/18/2021 14:21	SMPL
M2696-07	BGHJ6S	13082	2.7485 -		SFAM01.1	PPB	6/18/2021 14:23	SMPL
M2696-08	BGHJ7	2918	0.619 -		SFAM01.1	PPB	6/18/2021 14:26	SMPL
M2696-09	BGHJ9	6179	1.3022 -		SFAM01.1	PPB	6/18/2021 14:28	SMPL
M2696-10	BGHK1	1678	0.3592 -		SFAM01.1	PPB	6/18/2021 14:30	SMPL
M2696-11	BGHK2	938	0.2041 -		SFAM01.1	PPB	6/18/2021 14:32	SMPL
M2696-12	BGHK3	11927	2.5065 -		SFAM01.1	PPB	6/18/2021 14:35	SMPL
CCV033	CCV033	24106	5.0582 -		SFAM01.1	PPB	6/18/2021 14:37	SMPL
CCB033	CCB033	-109	-0.0152 -		SFAM01.1	PPB	6/18/2021 14:39	SMPL