

LB 115024


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51825



A= 0.0000e+000

B= 1.9149e-004

C= 4.9663e-003

Rho= 0.9998916

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	0.007	0.007	12	0.000	12					1
0.2	0.200	0.198	-0.002	1009	0.0 %	1009					-1
2.5	2.500	2.459	-0.041	12815	0.0 %	12815					-2
5.0	5.000	5.002	0.002	26094	0.0 %	26094					0
7.5	7.500	7.605	0.105	39686	0.0 %	39686					1
10.0	10.000	9.929	-0.071	51825	0.0 %	51825					-1

LB115024 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 18 Jun 2021 15:03:08					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7471B	PPB	18 Jun 2021 15:05:17	S	Std
0.2 - 1	502	1009	-	0.2000	7471B	PPB	18 Jun 2021 15:07:34	S	Std
2.5 - 1	525	12815	-	2.5000	7471B	PPB	18 Jun 2021 15:09:50	S	Std
5.0 - 1	55	26094	-	5.0000	7471B	PPB	18 Jun 2021 15:12:07	S	Std
7.5 - 1	57.1	39686	-	7.5000	7471B	PPB	18 Jun 2021 15:14:23	S	Std
10.0 - 1	510	51825	-	10.0000	7471B	PPB	18 Jun 2021 15:16:39	S	Std
ICV69 - 1	ICV69	20543	3.9388	-	7471B	PPB	18 Jun 2021 15:19:23	U	SMPL
ICB69 - 1	ICB69	-151	-0.0239	-	7471B	PPB	18 Jun 2021 15:21:38	U	SMPL
CCV34 - 1	CCV34	26565	5.0920	-	7471B	PPB	18 Jun 2021 15:23:54	U	SMPL
CCB34 - 1	CCB34	-168	-0.0272	-	7471B	PPB	18 Jun 2021 15:26:08	U	SMPL
CRA - 1	CRA	966	0.1899	-	7471B	PPB	18 Jun 2021 15:28:24	U	SMPL
HighStd - 1	HighStd	51986	9.9600	-	7471B	PPB	18 Jun 2021 15:30:39	U	SMPL
ChkStd - 1	ChkStd	36291	6.9545	-	7471B	PPB	18 Jun 2021 15:32:54	U	SMPL
PB137212BL - 1	PBS	-229	-0.0389	-	7471B	PPB	18 Jun 2021 15:35:12	U	SMPL
PB137212BS - 1	LCSS	18422	3.5327	-	7471B	PPB	18 Jun 2021 15:40:08	U	SMPL
M2759-01 - 1	TR-05-061721	5117	0.9848	-	7471B	PPB	18 Jun 2021 15:42:24	U	SMPL
M2764-01 - 1	PHC-250EM-001	2387	0.4621	-	7471B	PPB	18 Jun 2021 15:44:41	U	SMPL
M2764-03 - 1	PHC-250EM-002	698	0.1386	-	7471B	PPB	18 Jun 2021 15:46:57	U	SMPL
M2768-01 - 1	OK-01-061721	1288	0.2516	-	7471B	PPB	18 Jun 2021 15:49:15	U	SMPL
M2771-01 - 1	0338-COMPOSITE-DC-21	27638	5.2975	-	7471B	PPB	18 Jun 2021 15:51:30	U	SMPL
CCV35 - 1	CCV35	26402	5.0608	-	7471B	PPB	18 Jun 2021 15:53:45	U	SMPL
CCB35 - 1	CCB35	-120	-0.0180	-	7471B	PPB	18 Jun 2021 15:56:01	U	SMPL
M2772-01 - 1	SB-2	4248	0.8184	-	7471B	PPB	18 Jun 2021 16:03:30	U	SMPL
M2772-01DUP - 1	SB-2DUP	4397	0.8470	-	7471B	PPB	18 Jun 2021 16:05:45	U	SMPL
M2772-01MS - 1	SB-2MS	23890	4.5798	-	7471B	PPB	18 Jun 2021 16:08:00	U	SMPL
M2772-01MSD - 1	SB-2MSD	23472	4.4997	-	7471B	PPB	18 Jun 2021 16:12:53	U	SMPL
M2772-02 - 1	SB-12	5070	0.9758	-	7471B	PPB	18 Jun 2021 16:15:08	U	SMPL
M2772-03 - 1	SB-11	6319	1.2150	-	7471B	PPB	18 Jun 2021 16:17:24	U	SMPL
M2772-04 - 1	SB-1	979	0.1924	-	7471B	PPB	18 Jun 2021 16:19:40	U	SMPL
PB137213BL - 1	PBS	0	0.0050	-	7471B	PPB	18 Jun 2021 16:21:57	U	SMPL
PB137213BS - 1	LCSS	20044	3.8433	-	7471B	PPB	18 Jun 2021 16:24:14	U	SMPL
M2773-01 - 1	SB-1	4245	0.8179	-	7471B	PPB	18 Jun 2021 16:26:29	U	SMPL
CCV36 - 1	CCV36	26740	5.1255	-	7471B	PPB	18 Jun 2021 16:28:45	U	SMPL
CCB36 - 1	CCB36	-136	-0.0211	-	7471B	PPB	18 Jun 2021 16:31:00	U	SMPL
M2773-02 - 1	SB-2	52277	10.0157	-	7471B	PPB	18 Jun 2021 16:33:16	U	SMPL
M2773-03 - 1	SB-3	3771	0.7271	-	7471B	PPB	18 Jun 2021 16:35:31	U	SMPL
M2773-04 - 1	SB-4	23769	4.5566	-	7471B	PPB	18 Jun 2021 16:37:48	U	SMPL
M2773-05 - 1	SB-5	12245	2.3498	-	7471B	PPB	18 Jun 2021 16:40:04	U	SMPL
M2773-06 - 1	SB-6	9269	1.7799	-	7471B	PPB	18 Jun 2021 16:42:19	U	SMPL
M2773-07 - 1	SB-7	2357	0.4563	-	7471B	PPB	18 Jun 2021 16:44:35	U	SMPL
M2773-08 - 1	SB-8	2624	0.5074	-	7471B	PPB	18 Jun 2021 16:46:52	U	SMPL
M2773-09 - 1	SB-9	823	0.1626	-	7471B	PPB	18 Jun 2021 16:49:08	U	SMPL
M2773-10 - 1	SB-10	2918	0.5637	-	7471B	PPB	18 Jun 2021 16:51:25	U	SMPL
M2773-11 - 1	SB-9-AOC-2	10182	1.9548	-	7471B	PPB	18 Jun 2021 16:53:41	U	SMPL
CCV37 - 1	CCV37	26679	5.1138	-	7471B	PPB	18 Jun 2021 16:55:56	U	SMPL
CCB37 - 1	CCB37	-126	-0.0192	-	7471B	PPB	18 Jun 2021 16:58:11	U	SMPL
M2773-12 - 1	SB-6-AOC-2	5711	1.0986	-	7471B	PPB	18 Jun 2021 17:00:27	U	SMPL
M2773-12DUP - 1	SB-6-AOC-2DUP	6355	1.2219	-	7471B	PPB	18 Jun 2021 17:02:43	U	SMPL
M2773-12MS - 1	SB-6-AOC-2MS	25115	4.8143	-	7471B	PPB	18 Jun 2021 17:04:58	U	SMPL
M2773-12MSD - 1	SB-6-AOC-2MSD	26792	5.1355	-	7471B	PPB	18 Jun 2021 17:07:13	U	SMPL
M2772-01LX5 - 1		763	0.1511	-	7471B	PPB	18 Jun 2021 17:09:29	U	SMPL
M2772-01A - 1		25252	4.8406	-	7471B	PPB	18 Jun 2021 17:11:45	U	SMPL
M2773-12LX5 - 1		987	0.1940	-	7471B	PPB	18 Jun 2021 17:14:01	U	SMPL
M2773-12A - 1		26357	5.0522	-	7471B	PPB	18 Jun 2021 17:16:18	U	SMPL
M2773-02DLX2 - 1		26041	4.9917	-	7471B	PPB	18 Jun 2021 17:18:35	U	SMPL
CCV38 - 1	CCV38	26695	5.1169	-	7471B	PPB	18 Jun 2021 17:20:51	U	SMPL
CCB38 - 1	CCB38	-120	-0.0180	-	7471B	PPB	18 Jun 2021 17:23:06	U	SMPL