



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115025

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 8                 | 27.00      | 80.6      | 80.6            | 06/18/2021 | 09:50     |
| 2   | M2762-02    |               | 1  | 9                 | 100.00     | >212.0    | >212.0          | 06/18/2021 | 10:20     |
| 3   | M2762-02DUP |               | 1  | 9                 | 100.00     | >212.0    | >212.0          | 06/18/2021 | 10:45     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB 11 5025

WorkList Name : fp61821      WorkList ID : 149948      Department : Wet-Chemistry      Date : 06-18-2021 09:19:22

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 06/24/2021 | Water  | M2762-02 | Flash Point | Cool 4 deg C | ROYF02   | F52                         | WC_AQUEOUS_DBD  | 06/17/2021   | 1010A  |

Date/Time 06/18/21 09:25  
 Raw Sample Received by: KS (Webber)  
 Raw Sample Relinquished by: JR (Stu)

Date/Time 06/18/21 10:50  
 Raw Sample Received by: JR (Stu)  
 Raw Sample Relinquished by: KS (Webber)