


LB115035

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5056e-004

C= -3.3357e-003

Rho= 0.9995450

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.003	-0.003	0	0.000	0					-
0.2	0.200	0.185	-0.015	752	0.0 %	752					-8
2.5	2.500	2.504	0.004	10008	0.0 %	10008					0
5.0	5.000	4.933	-0.067	19701	0.0 %	19701					-1
7.5	7.500	7.723	0.223	30834	0.0 %	30834					3
10.0	10.000	9.858	-0.142	39358	0.0 %	39358					-1

LB115035 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 21 Jun 2021 11:19:12					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7470A	PPB	21 Jun 2021 11:24:35	S	Std
0.2 - 1	50.2	752	-	0.2000	7470A	PPB	21 Jun 2021 11:26:52	S	Std
2.5 - 1	52.5	10008	-	2.5000	7470A	PPB	21 Jun 2021 11:29:09	S	Std
5.0 - 1	55	19701	-	5.0000	7470A	PPB	21 Jun 2021 11:31:25	S	Std
7.5 - 1	57.5	30834	-	7.5000	7470A	PPB	21 Jun 2021 11:33:41	S	Std
10.0 - 1	57.5	39358	-	10.0000	7470A	PPB	21 Jun 2021 11:41:05	S	Std
ICV70 - 1	ICV70	15384	3.8513	-	7470A	PPB	21 Jun 2021 11:43:48	U	SMPL
ICB70 - 1	ICB70	-152	-0.0414	-	7470A	PPB	21 Jun 2021 11:46:03	U	SMPL
CCV39 - 1	CCV39	19434	4.8661	-	7470A	PPB	21 Jun 2021 11:48:18	U	SMPL
CCB39 - 1	CCB39	-222	-0.0590	-	7470A	PPB	21 Jun 2021 11:50:32	U	SMPL
CRA - 1	CRA	702	0.1726	-	7470A	PPB	21 Jun 2021 11:52:48	U	SMPL
HighStd - 1	HighStd	39310	9.8463	-	7470A	PPB	21 Jun 2021 11:55:02	U	SMPL
ChkStd - 1	ChkStd	29035	7.2718	-	7470A	PPB	21 Jun 2021 11:57:18	U	SMPL
PB137210BL - 1	PBW	-220	-0.0585	-	7470A	PPB	21 Jun 2021 11:59:35	U	SMPL
PB137210BS - 1	LCSW	15509	3.8827	-	7470A	PPB	21 Jun 2021 12:01:51	U	SMPL
M2762-01 - 1	WC_SOLIDS_DBD	-133	-0.0367	-	7470A	PPB	21 Jun 2021 12:04:07	U	SMPL
M2762-01DUP - 1	WC_SOLIDS_DBDDUP	-25	-0.0096	-	7470A	PPB	21 Jun 2021 12:06:24	U	SMPL
M2762-01MS - 1	WC_SOLIDS_DBDMS	15577	3.8997	-	7470A	PPB	21 Jun 2021 12:11:12	U	SMPL
M2762-01MSD - 1	WC_SOLIDS_DBDMSD	15626	3.9120	-	7470A	PPB	21 Jun 2021 12:13:29	U	SMPL
M2764-02 - 1	PHC-250EM-001	-88	-0.0254	-	7470A	PPB	21 Jun 2021 12:15:45	U	SMPL
CCV40 - 1	CCV40	19705	4.9340	-	7470A	PPB	21 Jun 2021 12:18:00	U	SMPL
CCB40 - 1	CCB40	-164	-0.0444	-	7470A	PPB	21 Jun 2021 12:20:15	U	SMPL
M2764-04 - 1	PHC-250EM-002	-2	-0.0038	-	7470A	PPB	21 Jun 2021 12:22:29	U	SMPL
PB137211BL - 1	PBW	-12	-0.0063	-	7470A	PPB	21 Jun 2021 12:24:44	U	SMPL
PB137211BS - 1	LCSW	13358	3.3437	-	7470A	PPB	21 Jun 2021 12:26:59	U	SMPL
M2741-01 - 1	0337-FEILD-BLANK	-105	-0.0296	-	7470A	PPB	21 Jun 2021 12:29:15	U	SMPL
M2771-02 - 1	0370-COMPOSITE-DC-21-MOD-ELU	120	0.0267	-	7470A	PPB	21 Jun 2021 12:31:30	U	SMPL
M2771-02DUP - 1	0370-COMPOSITE-DC-21-MOD-ELUDUP	112	0.0247	-	7470A	PPB	21 Jun 2021 12:33:48	U	SMPL
M2771-02MS - 1	0370-COMPOSITE-DC-21-MOD-ELUMS	15134	3.7887	-	7470A	PPB	21 Jun 2021 12:36:05	U	SMPL
M2771-02MSD - 1	0370-COMPOSITE-DC-21-MOD-ELUMSD	14752	3.6930	-	7470A	PPB	21 Jun 2021 12:38:22	U	SMPL
M2771-03 - 1	0370-COMPOSITE-DC-21-MOD-ELU	-23	-0.0091	-	7470A	PPB	21 Jun 2021 12:40:40	U	SMPL
CCV41 - 1	CCV41	19892	4.9809	-	7470A	PPB	21 Jun 2021 12:42:56	U	SMPL
CCB41 - 1	CCB41	-167	-0.0452	-	7470A	PPB	21 Jun 2021 12:45:11	U	SMPL
PB137210TB - 1	PB137181TB	-41	-0.0136	-	7470A	PPB	21 Jun 2021 12:47:27	U	SMPL
M2762-01LX5 - 1		-23	-0.0091	-	7470A	PPB	21 Jun 2021 12:49:42	U	SMPL
M2762-01A - 1		15415	3.8591	-	7470A	PPB	21 Jun 2021 12:51:58	U	SMPL
M2771-02LX5 - 1		-84	-0.0244	-	7470A	PPB	21 Jun 2021 12:54:13	U	SMPL
M2771-02A - 1		15397	3.8546	-	7470A	PPB	21 Jun 2021 12:56:29	U	SMPL
CCV42 - 1	CCV42	19336	4.8416	-	7470A	PPB	21 Jun 2021 13:01:30	U	SMPL
CCB42 - 1	CCB42	-191	-0.0512	-	7470A	PPB	21 Jun 2021 13:03:45	U	SMPL