



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115044

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/22/2021	08:27
2	CAL2	1	Water	NA	NA	20.6	7.00	06/22/2021	08:28
3	CAL3	1	Water	NA	NA	20.6	10.02	06/22/2021	08:29
4	ICV	1	Water	NA	NA	20.6	6.99	06/22/2021	08:30
5	CCV1	1	Water	NA	NA	20.6	2.01	06/22/2021	08:32
6	M2786-02	1	Water	NA	NA	24.0	5.88	06/22/2021	08:35
7	M2786-04	1	Water	NA	NA	23.9	5.73	06/22/2021	08:36
8	M2786-06	1	Water	NA	NA	23.9	8.77	06/22/2021	08:39
9	M2786-08	1	Water	NA	NA	23.8	6.25	06/22/2021	08:40
10	M2786-10	1	Water	NA	NA	23.9	6.70	06/22/2021	08:43
11	M2787-01	1	Water	NA	NA	23.9	7.56	06/22/2021	08:45
12	M2787-02	1	Water	NA	NA	24.0	7.59	06/22/2021	08:46
13	M2787-02DUP	1	Water	NA	NA	24.1	7.60	06/22/2021	08:47
14	CCV2	1	Water	NA	NA	20.9	12.02	06/22/2021	08:49

115049

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2796 w WorkList ID : 149991 Department : Wet-Chemistry Date : 06-22-2021 08:23:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/21/2021	Water	M2786-02	pH	Cool 4 deg C	PSEG03	J11	33249	06/21/2021	9040C
06/21/2021	Water	M2786-04	pH	Cool 4 deg C	PSEG03	J11	33250	06/21/2021	9040C
06/21/2021	Water	M2786-06	pH	Cool 4 deg C	PSEG03	J11	33251	06/21/2021	9040C
06/21/2021	Water	M2786-08	pH	Cool 4 deg C	PSEG03	J11	33264	06/21/2021	9040C
06/21/2021	Water	M2786-10	pH	Cool 4 deg C	PSEG03	J11	33265	06/21/2021	9040C
06/21/2021	Water	M2787-01	pH	Cool 4 deg C	PSEG03	J11	41216	06/21/2021	9040C
06/21/2021	Water	M2787-02	pH	Cool 4 deg C	PSEG03	J11	41828	06/21/2021	9040C

Date/Time 06.22.21 08:25
 Raw Sample Received by: JO [Signature]
 Raw Sample Relinquished by: CP [Signature]

Date/Time 06.22.21 10:30
 Raw Sample Received by: CP [Signature]
 Raw Sample Relinquished by: JO [Signature]