



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115048

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	06/22/2021	11:20
2	M2786-02		1	8	100.00	>212.0	>212.0	06/22/2021	11:45
3	M2786-02DUP		1	8	100.00	>212.0	>212.0	06/22/2021	12:10
4	M2786-04		1	9	100.00	>212.0	>212.0	06/22/2021	12:35
5	M2786-06		1	8	100.00	>212.0	>212.0	06/22/2021	13:05
6	M2786-08		1	9	100.00	>212.0	>212.0	06/22/2021	13:35
7	M2786-10		1	8	100.00	>212.0	>212.0	06/22/2021	13:58
8	M2787-01		1	9	100.00	>212.0	>212.0	06/22/2021	14:20
9	M2787-02		1	9	100.00	>212.0	>212.0	06/22/2021	14:50
10	M2802-03		1	8	100.00	>212.0	>212.0	06/22/2021	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp62221

WorkList ID : 150009

Department : Wet-Chemistry

Date : 06-22-2021 10:20:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/28/2021	Water	M2786-02	Flash Point	Cool 4 deg C	PSEG03	J11	33249	06/21/2021	1010A
06/28/2021	Water	M2786-04	Flash Point	Cool 4 deg C	PSEG03	J11	33250	06/21/2021	1010A
06/28/2021	Water	M2786-06	Flash Point	Cool 4 deg C	PSEG03	J11	33251	06/21/2021	1010A
06/28/2021	Water	M2786-08	Flash Point	Cool 4 deg C	PSEG03	J11	33264	06/21/2021	1010A
06/28/2021	Water	M2786-10	Flash Point	Cool 4 deg C	PSEG03	J11	33265	06/21/2021	1010A
06/28/2021	Water	M2787-01	Flash Point	Cool 4 deg C	PSEG03	J11	41216	06/21/2021	1010A
06/28/2021	Water	M2787-02	Flash Point	Cool 4 deg C	PSEG03	J11	41828	06/21/2021	1010A
06/29/2021	Water	M2802-03	Flash Point	Cool 4 deg C	PSEG03	J12	41175	06/22/2021	1010A

Date/Time 06/22/21 11:00
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: SP (at chem)

Date/Time 06/22/21 15:20
 Raw Sample Received by: SP (at chem)
 Raw Sample Relinquished by: KS (Wetchem)