

LB115049


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53699



A= 0.0000e+000

B= 1.8522e-004

C= -4.4945e-002

Rho= 0.9998505

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/20
0.0	0.000	-0.020	-0.020	135	0.000	135					-
0.2	0.200	0.190	-0.010	1266	0.0 %	1266					-5
2.5	2.500	2.472	-0.028	13591	0.0 %	13591					-1
5.0	5.000	5.056	0.056	27543	0.0 %	27543					-1
7.5	7.500	7.601	0.101	41280	0.0 %	41280					-1
10.0	10.000	9.901	-0.099	53699	0.0 %	53699					-1

LB115049 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 22 Jun 2021 14:19:43							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	S0	135	-	0.00007471B	PPB		22 Jun 2021 14:23:31	S	Std		
0.2 - 1	S0.2	1266	-	0.20007471B	PPB		22 Jun 2021 14:25:47	S	Std		
2.5 - 1	S2.5	13591	-	2.50007471B	PPB		22 Jun 2021 14:28:04	S	Std		
5.0 - 1	S5	27543	-	5.00007471B	PPB		22 Jun 2021 14:30:21	S	Std		
7.5 - 1	S7.5	41280	-	7.50007471B	PPB		22 Jun 2021 14:32:37	S	Std		
10.0 - 1	S10	53699	-	10.00007471B	PPB		22 Jun 2021 14:37:27	S	Std		
ICV72 - 1	ICV72	21020	3.8483	-7471B	PPB		22 Jun 2021 14:40:09	U	SMPL		
ICB72 - 1	ICB72	-33	-0.0511	-7471B	PPB		22 Jun 2021 14:42:24	U	SMPL		
CCV49 - 1	CCV49	27857	5.1146	-7471B	PPB		22 Jun 2021 14:44:39	U	SMPL		
CCB49 - 1	CCB49	-40	-0.0524	-7471B	PPB		22 Jun 2021 14:46:53	U	SMPL		
CRA - 1	CRA	1244	0.1855	-7471B	PPB		22 Jun 2021 14:49:09	U	SMPL		
HighStd - 1	HighStd	54341	10.0198	-7471B	PPB		22 Jun 2021 14:51:24	U	SMPL		
ChkStd - 1	ChkStd	38284	7.0458	-7471B	PPB		22 Jun 2021 14:53:39	U	SMPL		
PB137254BL - 1	PBS	-14	-0.0475	-7471B	PPB		22 Jun 2021 14:55:58	U	SMPL		
PB137254BS - 1	LCSS	21295	3.8992	-7471B	PPB		22 Jun 2021 14:58:15	U	SMPL		
M2776-01 - 1	0339-DISCRETE-CORE-1	15911	2.9020	-7471B	PPB		22 Jun 2021 15:00:31	U	SMPL		
M2778-07 - 1	CF	4156	0.7248	-7471B	PPB		22 Jun 2021 15:02:49	U	SMPL		
M2790-01 - 1	SU-04-062121	2109	0.3457	-7471B	PPB		22 Jun 2021 15:05:06	U	SMPL		
M2793-02 - 1	SB-2	408	0.0306	-7471B	PPB		22 Jun 2021 15:07:24	U	SMPL		
M2793-02DUP - 1	SB-2DUP	408	0.0306	-7471B	PPB		22 Jun 2021 15:09:40	U	SMPL		
CCV50 - 1	CCV50	27403	5.0305	-7471B	PPB		22 Jun 2021 15:11:58	U	SMPL		
CCB50 - 1	CCB50	-31	-0.0507	-7471B	PPB		22 Jun 2021 15:14:15	U	SMPL		
M2793-02MS - 1	SB-2MS	20834	3.8138	-7471B	PPB		22 Jun 2021 15:16:31	U	SMPL		
M2793-02MSD - 1	SB-2MSD	20133	3.6840	-7471B	PPB		22 Jun 2021 15:18:46	U	SMPL		
M2793-04 - 1	SB-4	361	0.0219	-7471B	PPB		22 Jun 2021 15:21:02	U	SMPL		
M2793-05 - 1	SB-5	280	0.0069	-7471B	PPB		22 Jun 2021 15:23:17	U	SMPL		
M2795-01 - 1	B1-0-2	81632	15.0745	-7471B	PPB		22 Jun 2021 15:25:32	U	SMPL		
M2795-02 - 1	B1-6-8	16674	3.0433	-7471B	PPB		22 Jun 2021 15:27:48	U	SMPL		
M2795-03 - 1	B2-0-2	75643	13.9653	-7471B	PPB		22 Jun 2021 15:30:07	U	SMPL		
M2795-04 - 1	B2-6-8	10585	1.9156	-7471B	PPB		22 Jun 2021 15:32:24	U	SMPL		
M2795-05 - 1	B3-0-2	169490	31.3472	-7471B	PPB		22 Jun 2021 15:34:42	U	SMPL		
M2795-06 - 1	B3-6-8	7928	1.4234	-7471B	PPB		22 Jun 2021 15:36:58	U	SMPL		
CCV51 - 1	CCV51	27570	5.0614	-7471B	PPB		22 Jun 2021 15:39:16	U	SMPL		
CCB51 - 1	CCB51	-60	-0.0561	-7471B	PPB		22 Jun 2021 15:41:31	U	SMPL		
M2795-07 - 1	B4-0-2	11415	2.0693	-7471B	PPB		22 Jun 2021 15:43:47	U	SMPL		
M2800-01 - 1	BORING-ADDITION	990	0.1384	-7471B	PPB		22 Jun 2021 15:46:02	U	SMPL		
M2801-01 - 1	S-1	2082	0.3407	-7471B	PPB		22 Jun 2021 15:48:18	U	SMPL		
M2793-02LX5 - 1		161	-0.0151	-7471B	PPB		22 Jun 2021 15:50:33	U	SMPL		
M2793-02A - 1		20690	3.7872	-7471B	PPB		22 Jun 2021 15:52:49	U	SMPL		
M2795-01DLX5 - 1		15959	2.9109	-7471B	PPB		22 Jun 2021 15:55:05	U	SMPL		
M2795-03DLX5 - 1		14838	2.7033	-7471B	PPB		22 Jun 2021 15:57:22	U	SMPL		
M2795-05DLX5 - 1		34893	6.4178	-7471B	PPB		22 Jun 2021 15:59:38	U	SMPL		
CCV52 - 1	CCV52	27817	5.1072	-7471B	PPB		22 Jun 2021 16:01:55	U	SMPL		
CCB52 - 1	CCB52	-70	-0.0579	-7471B	PPB		22 Jun 2021 16:04:11	U	SMPL		