



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115053

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/22/2021	15:35
2	CAL2	1	Water	NA	NA	20.6	7.00	06/22/2021	15:36
3	CAL3	1	Water	NA	NA	20.6	10.03	06/22/2021	15:37
4	ICV	1	Water	NA	NA	20.7	6.99	06/22/2021	15:39
5	CCV1	1	Water	NA	NA	20.8	2.01	06/22/2021	15:40
6	M2802-03	1	Water	NA	NA	24.0	7.32	06/22/2021	15:42
7	M2812-07	1	Water	NA	NA	20.9	8.34	06/22/2021	15:44
8	M2814-01	1	Water	NA	NA	21.7	5.09	06/22/2021	15:45
9	M2814-01DUP	1	Water	NA	NA	21.6	5.10	06/22/2021	15:46
10	CCV2	1	Water	NA	NA	20.5	12.02	06/22/2021	15:47

115053

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2812 WorkList ID : 150016 Department : Wet-Chemistry Date : 06-22-2021 15:25:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2021	Water	M2802-03	A pH	Cool 4 deg C	PSEG03	J12	41175	06/22/2021	9040C
06/22/2021	Water	M2812-07	E pH	Cool 4 deg C	PSEG03	J12	KG-WATER	06/22/2021	9040C
06/21/2021	Water	M2814-01	A pH	Cool 4 deg C	STER08	J21	POOL-WATER	06/21/2021	9040C

Date/Time 06/22/21 15:30
 Raw Sample Received by: JD out Chem
 Raw Sample Relinquished by: CP SM

Date/Time 06-22-21 14:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JD out Chem