



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115056

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/22/2021	17:35
2	CAL2	1	Water	NA	NA	20.6	7.00	06/22/2021	15:36
3	CAL3	1	Water	NA	NA	20.6	10.03	06/22/2021	15:37
4	ICV	1	Water	NA	NA	20.7	6.99	06/22/2021	15:39
5	CCV1	1	Water	NA	NA	20.8	2.01	06/22/2021	15:40
6	M2812-02	1	Solid	20.03	20	23.4	8.92	06/22/2021	15:50
7	M2812-02DUP	1	Solid	20.04	20	23.5	8.93	06/22/2021	15:51
8	M2812-05	1	Solid	20.02	20	23.3	8.34	06/22/2021	16:00
9	CCV2	1	Water	NA	NA	20.6	12.02	06/22/2021	16:01

2B115056

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2812 s WorkList ID : 150019 Department : Wet-Chemistry Date : 06-22-2021 15:28:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2021	Solid	M2812-02	Corrosivity	Cool 4 deg C	PSEG03	J12	356-349	06/22/2021	9045D
06/22/2021	Solid	M2812-05	Corrosivity	Cool 4 deg C	PSEG03	J12	293-347	06/22/2021	9045D

Date/Time 06/22/21 15:20
 Raw Sample Received by: JB Clut Chem
 Raw Sample Relinquished by: CP SM

Date/Time 06/22/21 17:40
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JB (wet Chem)