


LB115059
INSTRUMENT ID: CV1

7470A



A= 0.0000e+000

B= 2.4869e-004

C= 8.3283e-003

Rho= 0.9996655

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.0	0.000	0.018	0.018	39	0.000	39					-
0.2	0.200	0.204	0.004	788	0.0 %	788					2
2.5	2.500	2.410	-0.090	9656	0.0 %	9656					-4
5.0	5.000	5.014	0.014	20127	0.0 %	20127					0
7.5	7.500	7.676	0.176	30833	0.0 %	30833					2
10.0	10.000	9.878	-0.122	39686	0.0 %	39686					-1

LB115059 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jun 2021 10:28:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7470A	PPB	23 Jun 2021 10:34:22	S	Std
0.2 - 1	50.2	788	-	0.2000	7470A	PPB	23 Jun 2021 10:36:39	S	Std
2.5 - 1	52.5	9656	-	2.5000	7470A	PPB	23 Jun 2021 10:38:55	S	Std
5.0 - 1	55	20127	-	5.0000	7470A	PPB	23 Jun 2021 10:41:12	S	Std
7.5 - 1	57.5	30833	-	7.5000	7470A	PPB	23 Jun 2021 10:43:28	S	Std
10.0 - 1	510	39686	-	10.0000	7470A	PPB	23 Jun 2021 10:45:44	S	Std
ICV73 - 1	ICV73	15072	3.7566	-	7470A	PPB	23 Jun 2021 10:48:27	U	SMPL
ICB73 - 1	ICB73	-122	-0.0220	-	7470A	PPB	23 Jun 2021 10:50:42	U	SMPL
CCV53 - 1	CCV53	20135	5.0158	-	7470A	PPB	23 Jun 2021 10:55:40	U	SMPL
CCB53 - 1	CCB53	-105	-0.0178	-	7470A	PPB	23 Jun 2021 10:57:55	U	SMPL
CRA - 1	CRA	775	0.2011	-	7470A	PPB	23 Jun 2021 11:00:10	U	SMPL
HighStd - 1	HighStd	39787	9.9031	-	7470A	PPB	23 Jun 2021 11:02:25	U	SMPL
ChkStd - 1	ChkStd	28003	6.9725	-	7470A	PPB	23 Jun 2021 11:04:40	U	SMPL
PB137273BL - 1	PBW	-192	-0.0394	-	7470A	PPB	23 Jun 2021 11:06:56	U	SMPL
PB137273BS - 1	LCSW	14094	3.5134	-	7470A	PPB	23 Jun 2021 11:09:13	U	SMPL
M2795-08 - 1	B1GW	16347	4.0737	-	7470A	PPB	23 Jun 2021 11:11:29	U	SMPL
M2795-08DUP - 1	B1GWDUP	17096	4.2600	-	7470A	PPB	23 Jun 2021 11:13:45	U	SMPL
M2795-08MS - 1	B1GWMS	30983	7.7136	-	7470A	PPB	23 Jun 2021 11:16:02	U	SMPL
M2795-08MSD - 1	B1GWMSD	31849	7.9289	-	7470A	PPB	23 Jun 2021 11:18:18	U	SMPL
M2795-09 - 1	B3GW	9956	2.4843	-	7470A	PPB	23 Jun 2021 11:20:34	U	SMPL
CCV54 - 1	CCV54	19697	4.9068	-	7470A	PPB	23 Jun 2021 11:22:50	U	SMPL
CCB54 - 1	CCB54	-90	-0.0141	-	7470A	PPB	23 Jun 2021 11:25:05	U	SMPL
M2795-10 - 1	B4GW	1545	0.3926	-	7470A	PPB	23 Jun 2021 11:27:21	U	SMPL
M2795-13 - 1	B1GW	16	0.0123	-	7470A	PPB	23 Jun 2021 11:29:36	U	SMPL
M2795-14 - 1	B3GW	46	0.0198	-	7470A	PPB	23 Jun 2021 11:31:51	U	SMPL
M2795-15 - 1	B4GW	34	0.0168	-	7470A	PPB	23 Jun 2021 11:34:06	U	SMPL
M2799-02 - 1	SB-1-GW	203	0.0588	-	7470A	PPB	23 Jun 2021 11:36:22	U	SMPL
M2799-03 - 1	SB-2-GW	120	0.0382	-	7470A	PPB	23 Jun 2021 11:38:39	U	SMPL
M2799-04 - 1	MW-8	65	0.0245	-	7470A	PPB	23 Jun 2021 11:40:55	U	SMPL
M2799-05 - 1	MW-8D	27	0.0150	-	7470A	PPB	23 Jun 2021 11:43:12	U	SMPL
M2812-07 - 1	KG-WATER	16438	4.0963	-	7470A	PPB	23 Jun 2021 11:45:29	U	SMPL
CCV55 - 1	CCV55	20089	5.0043	-	7470A	PPB	23 Jun 2021 11:47:44	U	SMPL
CCB55 - 1	CCB55	-113	-0.0198	-	7470A	PPB	23 Jun 2021 11:50:03	U	SMPL
M2795-08LX5 - 1		2951	0.7422	-	7470A	PPB	23 Jun 2021 11:52:18	U	SMPL
M2795-08A - 1		32925	8.1965	-	7470A	PPB	23 Jun 2021 11:54:33	U	SMPL
CCV56 - 1	CCV56	19610	4.8852	-	7470A	PPB	23 Jun 2021 11:56:48	U	SMPL
CCB56 - 1	CCB56	-110	-0.0190	-	7470A	PPB	23 Jun 2021 11:59:04	U	SMPL