


LB115062

7471B

INSTRUMENT ID: CV1

 μ Abs.:

54929



Linear

A= 0.0000e+000

B= 1.8290e-004

C= -9.6367e-004

Rho= 0.9998419

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	10	0.000	10					-
0.2	0.200	0.197	-0.003	1084	0.0 %	1084					-2
2.5	2.500	2.484	-0.016	13584	0.0 %	13584					-1
5.0	5.000	5.094	0.094	27857	0.0 %	27857					2
7.5	7.500	7.379	-0.121	40350	0.0 %	40350					-2
10.0	10.000	10.045	0.045	54929	0.0 %	54929					0

LB115062 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 23 Jun 2021 12:44:15					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7471B	PPB	23 Jun 2021 13:26:53	S	Std
0.2 - 1	50.2	1084	-	0.2000	7471B	PPB	23 Jun 2021 13:29:09	S	Std
2.5 - 1	52.5	13584	-	2.5000	7471B	PPB	23 Jun 2021 13:31:26	S	Std
5.0 - 1	55	27857	-	5.0000	7471B	PPB	23 Jun 2021 13:33:43	S	Std
7.5 - 1	57.5	40350	-	7.5000	7471B	PPB	23 Jun 2021 13:35:59	S	Std
10.0 - 1	510	54929	-	10.0000	7471B	PPB	23 Jun 2021 13:38:16	S	Std
ICV75 - 1	ICV75	21762	3.9792	-	7471B	PPB	23 Jun 2021 13:41:03	U	SMPL
ICB75 - 1	ICB75	-113	-0.0216	-	7471B	PPB	23 Jun 2021 13:43:18	U	SMPL
CCV57 - 1	CCV57	27663	5.0585	-	7471B	PPB	23 Jun 2021 13:45:33	U	SMPL
CCB57 - 1	CCB57	-132	-0.0251	-	7471B	PPB	23 Jun 2021 13:47:48	U	SMPL
CRA - 1	CRA	1075	0.1957	-	7471B	PPB	23 Jun 2021 13:50:04	U	SMPL
HighStd - 1	HighStd	55196	10.0942	-	7471B	PPB	23 Jun 2021 13:52:19	U	SMPL
ChkStd - 1	ChkStd	37081	6.7810	-	7471B	PPB	23 Jun 2021 13:54:33	U	SMPL
PB137285BL - 1	PBS	-194	-0.0364	-	7471B	PPB	23 Jun 2021 13:56:52	U	SMPL
PB137285BS - 1	LCSS	20010	3.6588	-	7471B	PPB	23 Jun 2021 13:59:09	U	SMPL
M2802-01 - 1	30759	527	0.0954	-	7471B	PPB	23 Jun 2021 14:01:26	U	SMPL
M2802-02 - 1	41817	244	0.0437	-	7471B	PPB	23 Jun 2021 14:03:41	U	SMPL
M2810-01 - 1	NB-08-062221	2725	0.4974	-	7471B	PPB	23 Jun 2021 14:05:58	U	SMPL
M2811-01 - 1	HR-04-062221	2015	0.3676	-	7471B	PPB	23 Jun 2021 14:08:15	U	SMPL
M2812-01 - 1	356-349	29778	5.4453	-	7471B	PPB	23 Jun 2021 14:10:30	U	SMPL
CCV58 - 1	CCV58	27363	5.0036	-	7471B	PPB	23 Jun 2021 14:12:45	U	SMPL
CCB58 - 1	CCB58	-145	-0.0275	-	7471B	PPB	23 Jun 2021 14:15:01	U	SMPL
M2812-04 - 1	293-347	25807	4.7191	-	7471B	PPB	23 Jun 2021 14:17:17	U	SMPL
M2815-01 - 1	SH-1	2570	0.4691	-	7471B	PPB	23 Jun 2021 14:19:32	U	SMPL
M2815-01DUP - 1	SH-1DUP	2496	0.4555	-	7471B	PPB	23 Jun 2021 14:21:47	U	SMPL
M2815-01MS - 1	SH-1MS	22695	4.1499	-	7471B	PPB	23 Jun 2021 14:24:03	U	SMPL
M2815-01MSD - 1	SH-1MSD	21755	3.9780	-	7471B	PPB	23 Jun 2021 14:26:18	U	SMPL
M2816-01 - 1	OK-02-062221	1288	0.2346	-	7471B	PPB	23 Jun 2021 14:28:35	U	SMPL
M2815-01LX5 - 1		454	0.0821	-	7471B	PPB	23 Jun 2021 14:30:51	U	SMPL
M2815-01A - 1		24373	4.4568	-	7471B	PPB	23 Jun 2021 14:33:08	U	SMPL
CCV59 - 1	CCV59	27558	5.0393	-	7471B	PPB	23 Jun 2021 14:35:25	U	SMPL
CCB59 - 1	CCB59	-161	-0.0304	-	7471B	PPB	23 Jun 2021 14:37:40	U	SMPL