


LB115-069

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5194e-004

C= 1.5767e-002

Rho= 0.9999592

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	13	0.000	13					
0.2	0.200	0.205	0.005	751	0.0 %	751					
2.5	2.500	2.483	-0.017	9791	0.0 %	9791					
5.0	5.000	4.951	-0.049	19588	0.0 %	19588					
7.5	7.500	7.558	0.058	29935	0.0 %	29935					
10.0	10.000	9.985	-0.015	39569	0.0 %	39569					

13
751
9791
19588
29935
39569

LB115069 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Jun 2021 15:06:24					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	7470A	PPB	23 Jun 2021 15:08:39	S	Std
0.2 - 1	50.2	751	-	0.2000	7470A	PPB	23 Jun 2021 15:10:56	S	Std
2.5 - 1	52.5	9791	-	2.5000	7470A	PPB	23 Jun 2021 15:13:12	S	Std
5.0 - 1	55	19588	-	5.0000	7470A	PPB	23 Jun 2021 15:15:29	S	Std
7.5 - 1	57.5	29935	-	7.5000	7470A	PPB	23 Jun 2021 15:17:45	S	Std
10.0 - 1	510	39569	-	10.0000	7470A	PPB	23 Jun 2021 15:22:41	S	Std
ICV76 - 1	ICV76	16126	4.0786	-	7470A	PPB	23 Jun 2021 15:27:55	U	SMPL
ICB76 - 1	ICB76	-137	-0.0187	-	7470A	PPB	23 Jun 2021 15:30:09	U	SMPL
CCV60 - 1	CCV60	19242	4.8637	-	7470A	PPB	23 Jun 2021 15:32:23	U	SMPL
CCB60 - 1	CCB60	-123	-0.0152	-	7470A	PPB	23 Jun 2021 15:34:38	U	SMPL
CRA - 1	CRA	727	0.1989	-	7470A	PPB	23 Jun 2021 15:36:53	U	SMPL
HighStd - 1	HighStd	38791	9.7889	-	7470A	PPB	23 Jun 2021 15:39:08	U	SMPL
ChkStd - 1	ChkStd	27127	6.8502	-	7470A	PPB	23 Jun 2021 15:41:23	U	SMPL
PB137300BL - 1	PBW	-185	-0.0308	-	7470A	PPB	23 Jun 2021 15:43:40	U	SMPL
PB137300BS - 1	LCSW	17825	4.5067	-	7470A	PPB	23 Jun 2021 15:45:56	U	SMPL
M2812-08 - 1	KG-WATER	-79	-0.0041	-	7470A	PPB	23 Jun 2021 15:48:12	U	SMPL
M2812-08DUP - 1	KG-WATERDUP	24	0.0218	-	7470A	PPB	23 Jun 2021 15:50:28	U	SMPL
M2812-08MS - 1	KG-WATERMS	17608	4.4520	-	7470A	PPB	23 Jun 2021 15:52:45	U	SMPL
M2812-08MSD - 1	KG-WATERMSD	18031	4.5586	-	7470A	PPB	23 Jun 2021 15:55:01	U	SMPL
CCV61 - 1	CCV61	19528	4.9357	-	7470A	PPB	23 Jun 2021 15:57:17	U	SMPL
CCB61 - 1	CCB61	-146	-0.0210	-	7470A	PPB	23 Jun 2021 15:59:32	U	SMPL
PB137300TB - 1	PB1374277TB	-9	0.0135	-	7470A	PPB	23 Jun 2021 16:01:47	U	SMPL
M2812-08LX5 - 1		-4	0.0148	-	7470A	PPB	23 Jun 2021 16:04:03	U	SMPL
M2812-08A - 1		16614	4.2016	-	7470A	PPB	23 Jun 2021 16:06:18	U	SMPL
CCV62 - 1	CCV62	19372	4.8964	-	7470A	PPB	23 Jun 2021 16:08:33	U	SMPL
CCB62 - 1	CCB62	-109	-0.0117	-	7470A	PPB	23 Jun 2021 16:10:49	U	SMPL