



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115070

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/23/2021	15:36
2	CAL2	1	Water	NA	NA	20.3	7.00	06/23/2021	15:37
3	CAL3	1	Water	NA	NA	20.3	10.02	06/23/2021	15:38
4	ICV	1	Water	NA	NA	20.3	6.99	06/23/2021	15:40
5	CCV1	1	Water	NA	NA	20.4	2.01	06/23/2021	15:41
6	M2824-02	1	Solid	20.02	20	20.2	12.44	06/23/2021	15:44
7	M2835-01	1	Solid	20.03	20	20.5	7.13	06/23/2021	15:45
8	M2835-01DUP	1	Solid	20.04	20	20.6	7.14	06/23/2021	15:46
9	CCV2	1	Water	NA	NA	20.5	12.02	06/23/2021	15:47

VB 115070

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2835 WorkList ID : 150050 Department : Wet-Chemistry Date : 06-23-2021 15:34:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/23/2021	Solid	M2824-02	pH	Cool 4 deg C	JPCL01	J22	ABN-1	06/23/2021	9045D
06/23/2021	Solid	M2835-01	pH	Cool 4 deg C	TULL02	J31	FK-PS-01-075	06/23/2021	9045D

Date/Time 06/23/21 5:35
 Raw Sample Received by: JP (at MRM)
 Raw Sample Relinquished by: CP SM

Date/Time 06/23/21 6:35
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (at MRM)