



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115085

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/24/2021	11:10
2	CAL2	1	Water	NA	NA	20.4	7.00	06/24/2021	11:11
3	CAL3	1	Water	NA	NA	20.4	10.03	06/24/2021	11:13
4	ICV	1	Water	NA	NA	20.4	6.99	06/24/2021	11:15
5	CCV1	1	Water	NA	NA	20.4	2.01	06/24/2021	11:16
6	M2824-03	1	Solid	20.02	20	20.2	12.44	06/24/2021	11:20
7	M2824-03DUP	1	Solid	20.03	20	20.3	12.45	06/24/2021	11:21
8	M2826-01	1	Solid	20.03	20	20.4	5.49	06/24/2021	11:23
9	M2829-07	1	Solid	20.04	20	21.1	9.02	06/24/2021	11:25
10	M2829-08	1	Solid	20.02	20	21.1	8.73	06/24/2021	11:26
11	M2829-09	1	Solid	20.03	20	20.5	8.70	06/24/2021	11:29
12	M2829-10	1	Solid	20.01	20	20.7	8.85	06/24/2021	11:30
13	M2829-11	1	Solid	20.02	20	20.5	8.91	06/24/2021	11:33
14	M2829-12	1	Solid	20.03	20	20.5	8.81	06/24/2021	11:35
15	M2832-02	1	Solid	20.03	20	20.7	7.61	06/24/2021	11:40
16	CCV2	1	Water	NA	NA	20.3	12.02	06/24/2021	11:42
17	M2852-02	1	Solid	20.05	20	20.2	8.60	06/24/2021	11:45
18	M2852-05	1	Solid	20.03	20	20.5	8.20	06/24/2021	11:46
19	CCV3	1	Water	NA	NA	20.1	2.02	06/24/2021	11:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2829 WorkList ID : 150070 Department : Wet-Chemistry Date : 06-24-2021 10:57:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/23/2021	Solid	M2824-03	Corrosivity	Cool 4 deg C	JPCLO1	J22	ABN-1	06/23/2021	9045D
06/23/2021	Solid	M2826-01	Corrosivity	Cool 4 deg C	STAN07	J22	SS-1(SLUDGE)	06/23/2021	9045D
06/23/2021	Solid	M2829-07	Corrosivity	Cool 4 deg C	WALS01	J23	15B-STOCKPILE-1	06/23/2021	9045D
06/23/2021	Solid	M2829-08	Corrosivity	Cool 4 deg C	WALS01	J23	15B-STOCKPILE-2A	06/23/2021	9045D
06/23/2021	Solid	M2829-09	Corrosivity	Cool 4 deg C	WALS01	J23	15B-STOCKPILE-2B	06/23/2021	9045D
06/23/2021	Solid	M2829-10	Corrosivity	Cool 4 deg C	WALS01	J23	15B-STOCKPILE-3A	06/23/2021	9045D
06/23/2021	Solid	M2829-11	Corrosivity	Cool 4 deg C	WALS01	J23	15B-STOCKPILE-3B	06/23/2021	9045D
06/23/2021	Solid	M2829-12	Corrosivity	Cool 4 deg C	WALS01	J23	15B-DUP-01	06/23/2021	9045D
06/23/2021	Solid	M2832-02	Corrosivity	Cool 4 deg C	PSEG03	J22	TP-3-WC	06/23/2021	9045D
06/23/2021	Solid	M2852-02	Corrosivity	Cool 4 deg C	PSEG03	J23	MANHOLE-WC-S	06/23/2021	9045D
06/23/2021	Solid	M2852-05	Corrosivity	Cool 4 deg C	PSEG03	J23	MANHOLE-WC-D	06/23/2021	9045D

Date/Time 06/24/21 11:00
 Raw Sample Received by: JO [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06/24/21 13:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]