



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115088

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/24/2021	16:24
2	CAL2	1	Water	NA	NA	20.3	7.00	06/24/2021	16:25
3	CAL3	1	Water	NA	NA	20.3	10.02	06/24/2021	16:26
4	ICV	1	Water	NA	NA	20.4	7.00	06/24/2021	16:27
5	CCV1	1	Water	NA	NA	20.4	2.01	06/24/2021	16:28
6	M2860-01	1	Solid	20.02	20	21.8	6.63	06/24/2021	16:29
7	M2860-01DUP	1	Solid	20.01	20	21.9	6.64	06/24/2021	16:30
8	CCV2	1	Water	NA	NA	20.3	12.02	06/24/2021	16:32

PH 115088

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2860 WorkList ID : 150086 Department : Wet-Chemistry Date : 06-24-2021 16:12:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/24/2021	Solid	M2860-01	pH	Cool 4 deg C	FREE02	J41	ASH	06/24/2021	9045D

Date/Time 06/24/21 16:20
Raw Sample Received by: JO GAT Chrm
Raw Sample Relinquished by: JO GAT Chrm

Date/Time 06/24/21 18:00
Raw Sample Received by: JO GAT Chrm
Raw Sample Relinquished by: JO GAT Chrm