



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115093

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/24/2021	16:24
2	CAL2	1	Water	NA	NA	20.3	7.00	06/24/2021	16:25
3	CAL3	1	Water	NA	NA	20.3	10.02	06/24/2021	16:26
4	ICV	1	Water	NA	NA	20.4	7.00	06/24/2021	16:27
5	CCV1	1	Water	NA	NA	20.4	2.01	06/24/2021	16:28
6	M2826-02	1	Water	NA	NA	20.1	8.35	06/24/2021	16:35
7	M2846-03	1	Water	NA	NA	20.1	6.92	06/24/2021	16:36
8	M2848-01	1	Water	NA	NA	21.6	6.71	06/24/2021	16:39
9	M2848-01DUP	1	Water	NA	NA	21.7	6.72	06/24/2021	16:40
10	CCV2	1	Water	NA	NA	20.4	12.02	06/24/2021	16:41

18715093

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2826 W WorkList ID : 150071 Department : Wet-Chemistry Date : 06-24-2021 10:58:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/23/2021	Water	M2826-02	Corrosivity	Cool 4 deg C	STAN07	J22	LS-1(LIQUID-BLUE)	06/23/2021	9040C
06/24/2021	Water	M2846-03	Corrosivity	Cool 4 deg C	PSEG03	J22	DRUM-1	06/24/2021	9040C
06/24/2021	Water	M2848-01	Corrosivity	Cool 4 deg C	PSEG03	J23	OILY-WATER-DRUM	06/24/2021	9040C

Date/Time 06/24/21 16:00
 Raw Sample Received by: JO GAT Chem
 Raw Sample Relinquished by: JO GAT Chem

Date/Time 06/24/21 18:00
 Raw Sample Received by: JO GAT Chem
 Raw Sample Relinquished by: JO GAT Chem