



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115094

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/24/2021	16:24
2	CAL2	1	Water	NA	NA	20.3	7.00	06/24/2021	16:25
3	CAL3	1	Water	NA	NA	20.3	10.02	06/24/2021	16:26
4	ICV	1	Water	NA	NA	20.4	7.00	06/24/2021	16:27
5	CCV1	1	Water	NA	NA	20.4	2.01	06/24/2021	16:28
6	M2845-01	1	Water	NA	NA	21.6	6.81	06/24/2021	17:38
7	M2845-02	1	Water	NA	NA	21.8	8.81	06/24/2021	17:40
8	M2847-02	1	Water	NA	NA	22.1	8.05	06/24/2021	17:42
9	M2847-03	1	Water	NA	NA	23.6	6.92	06/24/2021	17:44
10	M2847-04	1	Water	NA	NA	22.0	6.82	06/24/2021	17:45
11	M2847-05	1	Water	NA	NA	21.9	6.75	06/24/2021	17:46
12	M2847-06	1	Water	NA	NA	22.6	11.73	06/24/2021	17:48
13	M2849-01	1	Water	NA	NA	21.8	6.71	06/24/2021	17:50
14	M2849-01DUP	1	Water	NA	NA	21.9	6.72	06/24/2021	17:51
15	CCV2	1	Water	NA	NA	20.3	12.02	06/24/2021	18:00

VB 715094

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2847 W WorkList ID : 150093 Department : Wet-Chemistry Date : 06-24-2021 17:31:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/24/2021	Water	M2845-01	pH	Cool 4 deg C	PSEG03	J23	41234	06/24/2021	9040C
06/24/2021	Water	M2845-02	pH	Cool 4 deg C	PSEG03	J23	41235	06/24/2021	9040C
06/24/2021	Water	M2847-02	pH	Cool 4 deg C	PSEG03	J23	1928-29	06/24/2021	9040C
06/24/2021	Water	M2847-03	pH	Cool 4 deg C	PSEG03	J23	1970-WATER	06/24/2021	9040C
06/24/2021	Water	M2847-04	pH	Cool 4 deg C	PSEG03	J23	1962-1964	06/24/2021	9040C
06/24/2021	Water	M2847-05	pH	Cool 4 deg C	PSEG03	J23	1966-1969	06/24/2021	9040C
06/24/2021	Water	M2847-06	pH	Cool 4 deg C	PSEG03	J23	1965	06/24/2021	9040C
06/24/2021	Water	M2849-01	pH	Cool 4 deg C	PSEG03	J23	NWB-1613	06/24/2021	9040C

Date/Time 06/24/21 17:35
 Raw Sample Received by: SA GAT (nem)
 Raw Sample Relinquished by: SA GAT (nem)

Date/Time 06/24/21 18:35
 Raw Sample Received by: SA GAT (nem)
 Raw Sample Relinquished by: SA GAT (nem)