

LB115096


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

52944



A= 0.0000e+000

B= 1.8841e-004

C= 4.5151e-003

Rho= 0.9999656

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%CD
0.0	0.000	0.022	0.022	94	0.000	94					1
0.2	0.200	0.205	0.005	1066	0.0 %	1066					2
2.5	2.500	2.444	-0.056	12945	0.0 %	12945					-2
5.0	5.000	5.013	0.013	26581	0.0 %	26581					0
7.5	7.500	7.536	0.036	39976	0.0 %	39976					0
10.0	10.000	9.980	-0.020	52944	0.0 %	52944					0

LB115096 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Jun 2021 08:51:42

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	94	-	0.00007471B	PPB	25 Jun 2021 09:07:28	S	Std
0.2 - 1	50.2	1066	-	0.20007471B	PPB	25 Jun 2021 09:09:45	S	Std
2.5 - 1	52.5	12945	-	2.50007471B	PPB	25 Jun 2021 09:12:02	S	Std
5.0 - 1	55	26581	-	5.00007471B	PPB	25 Jun 2021 09:14:19	S	Std
7.5 - 1	57.5	39976	-	7.50007471B	PPB	25 Jun 2021 09:16:34	S	Std
10.0 - 1	59	52944	-	10.00007471B	PPB	25 Jun 2021 09:18:51	S	Std
ICV77 - 1	ICV77	21185	3.9960	-7471B	PPB	25 Jun 2021 09:21:42	U	SMPL
ICB77 - 1	ICB77	-59	-0.0066	-7471B	PPB	25 Jun 2021 09:23:57	U	SMPL
CCV63 - 1	CCV63	26723	5.0394	-7471B	PPB	25 Jun 2021 09:26:13	U	SMPL
CCB63 - 1	CCB63	-71	-0.0089	-7471B	PPB	25 Jun 2021 09:28:28	U	SMPL
CRA - 1	CRA	1027	0.1980	-7471B	PPB	25 Jun 2021 09:30:43	U	SMPL
HighStd - 1	HighStd	52939	9.9788	-7471B	PPB	25 Jun 2021 09:32:58	U	SMPL
ChkStd - 1	ChkStd	37342	7.0402	-7471B	PPB	25 Jun 2021 09:35:13	U	SMPL
PB137322BL - 1	PBS	-192	-0.0317	-7471B	PPB	25 Jun 2021 09:37:30	U	SMPL
PB137322BS - 1	LCSS	19718	3.7196	-7471B	PPB	25 Jun 2021 09:39:47	U	SMPL
M2819-01 - 1	CL-02-062321	4184	0.7928	-7471B	PPB	25 Jun 2021 09:47:06	U	SMPL
M2819-01DUP - 1	CL-02-062321DUP	4017	0.7614	-7471B	PPB	25 Jun 2021 09:49:22	U	SMPL
M2819-01MS - 1	CL-02-062321MS	22624	4.2671	-7471B	PPB	25 Jun 2021 09:51:39	U	SMPL
M2819-01MSD - 1	CL-02-062321MSD	23458	4.4243	-7471B	PPB	25 Jun 2021 09:53:56	U	SMPL
M2829-01 - 1	15B-STOCKPILE-1	25705	4.8476	-7471B	PPB	25 Jun 2021 09:56:11	U	SMPL
CCV64 - 1	CCV64	26078	4.9179	-7471B	PPB	25 Jun 2021 09:58:27	U	SMPL
CCB64 - 1	CCB64	-145	-0.0228	-7471B	PPB	25 Jun 2021 10:00:42	U	SMPL
M2829-02 - 1	15B-STOCKPILE-2A	24957	4.7067	-7471B	PPB	25 Jun 2021 10:02:58	U	SMPL
M2829-03 - 1	15B-STOCKPILE-2B	24190	4.5622	-7471B	PPB	25 Jun 2021 10:05:14	U	SMPL
M2829-04 - 1	15B-STOCKPILE-3A	12976	2.4493	-7471B	PPB	25 Jun 2021 10:07:30	U	SMPL
M2829-05 - 1	15B-STOCKPILE-3B	15568	2.9377	-7471B	PPB	25 Jun 2021 10:09:46	U	SMPL
M2829-06 - 1	15B-DUP-01	37136	7.0014	-7471B	PPB	25 Jun 2021 10:12:02	U	SMPL
M2832-01 - 1	TP-3-WC	3805	0.7214	-7471B	PPB	25 Jun 2021 10:14:18	U	SMPL
M2833-01 - 1	FK-BPM-01-D8-C9-D8A-C9A	853	0.1652	-7471B	PPB	25 Jun 2021 10:16:37	U	SMPL
M2834-01 - 1	FK-BPM-01-B8-C8-B8A-C8A	574	0.1127	-7471B	PPB	25 Jun 2021 10:18:53	U	SMPL
M2835-01 - 1	FK-PS-01-075	1902	0.3629	-7471B	PPB	25 Jun 2021 10:21:10	U	SMPL
M2842-01 - 1	BIN-2-72-11966-RB-21140	1472	0.2819	-7471B	PPB	25 Jun 2021 10:23:25	U	SMPL
CCV65 - 1	CCV65	26286	4.9571	-7471B	PPB	25 Jun 2021 10:25:41	U	SMPL
CCB65 - 1	CCB65	-131	-0.0202	-7471B	PPB	25 Jun 2021 10:27:56	U	SMPL
M2843-01 - 1	OR-02-062421	2024	0.3859	-7471B	PPB	25 Jun 2021 10:30:12	U	SMPL
M2846-01 - 1	1-4-FEET-TP	868	0.1681	-7471B	PPB	25 Jun 2021 10:32:28	U	SMPL
M2851-01 - 1	BACKFILL-ASHCROFT-REAR-07	163	0.0352	-7471B	PPB	25 Jun 2021 10:34:43	U	SMPL
M2851-03 - 1	BACKFILL-ASHCROFT-REAR-08	164	0.0354	-7471B	PPB	25 Jun 2021 10:36:59	U	SMPL
M2852-01 - 1	MANHOLE-WC-S	496	0.0980	-7471B	PPB	25 Jun 2021 10:39:14	U	SMPL
M2852-04 - 1	MANHOLE-WC-D	195	0.0413	-7471B	PPB	25 Jun 2021 10:41:30	U	SMPL
M2819-01LX5 - 1		761	0.1479	-7471B	PPB	25 Jun 2021 10:43:46	U	SMPL
M2819-01A - 1		23920	4.5113	-7471B	PPB	25 Jun 2021 10:46:03	U	SMPL
CCV66 - 1	CCV66	26547	5.0063	-7471B	PPB	25 Jun 2021 10:48:20	U	SMPL
CCB66 - 1	CCB66	-130	-0.0200	-7471B	PPB	25 Jun 2021 10:50:36	U	SMPL