


Clear table

Instrument ID: IC-2										Analyst :		AP		Method:		300.0	
ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst						
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-06-09_	6/9/21 10:00		AP/MA						
STD2	0.4050	0.6460	0.6420	2.1610	0.5510	0.9000	3.3620	2021-06-09_	6/9/21 10:31		AP/MA						
STD3	0.7770	1.1870	1.1850	3.9630	0.9920	1.9910	5.9610	2021-06-09_	6/9/21 11:01		AP/MA						
STD4	0.9810	1.4970	1.4970	5.0160	1.2490	2.5950	7.4670	2021-06-09_	6/9/21 11:32		AP/MA						
STD5	2.0500	2.9570	2.9600	9.8110	2.4440	5.0400	14.6060	2021-06-09_	6/9/21 12:02		AP/MA						
STD6	4.0040	5.9880	6.0010	19.9330	4.9770	10.0050	29.8620	2021-06-09_	6/9/21 12:33		AP/MA						
STD7	4.9840	7.5250	7.5140	25.1160	6.2870	12.4680	37.7420	2021-06-09_	6/9/21 13:04		AP/MA						
ICV	2.0300	2.8660	2.8700	9.5040	2.3710	5.0290	14.1730	2021-06-09_	6/9/21 13:34		AP/MA						
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-06-09_	6/9/21 14:05		AP/MA						
CCV	2.0640	2.9510	2.8480	9.7650	2.4040	4.5670	14.4960	2021-06-24_	6/24/21 10:07		AP/MA						
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-06-24_	6/24/21 10:43		AP/MA						
LB115097BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-06-24_	6/24/21 11:13		AP/MA						
LB115097BSW	2.0550	2.9450	2.8480	9.7550	2.4130	4.6760	14.4930	2021-06-24_	6/24/21 11:44		AP/MA						
M2846-03	0.1180	49.3110	0.0000	0.0000	0.2840	0.0000	3.2470	2021-06-24_	6/24/21 17:14		AP/MA						
M2846-03MS	2.5040	51.0050	1.1950	9.8500	2.5710	7.8100	17.0950	2021-06-24_	6/24/21 17:45		AP/MA						
M2846-03MSD	2.5100	51.2550	1.2600	10.0650	2.6910	8.3150	17.7950	2021-06-24_	6/24/21 18:15		AP/MA						
CCV	2.0570	2.9750	2.8730	9.7810	2.4350	5.4180	14.5670	2021-06-24_	6/24/21 19:47		AP/MA						
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-06-24_	6/24/21 20:18		AP/MA						

CALIBRATION OF COMPONENT F-

IC-2
6/15/2021
AM

Method: AnionsIC2-060921.mtw
Equation: $Q = 0.0792888 \cdot A + 0.819643$
RSD: 1.384 %
Correlation coefficient: 0.999896



K3 = 0 K2 = 0 K1 = 0.0792888 K0 = 0.819643

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	2.884	91.85	0.4	20	5.75		
3	6.111	185.7	0.8	20	5.75		
4	7.874	237	1	20	5.75		
5	17.13	506.8	2	20	5.75		
6	36.83	999.5	4	20	5.75		
7	44.32	1247	5	20	5.75		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-060921.mtw
 Equation: $Q = 0.117495 \cdot A + 1.62784$
 RSD: 1.057 %
 Correlation coefficient: 0.999939



K3 = 0 K2 = 0 K1 = 0.117495 K0 = 1.62784

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.479	96.04	0.6	20	8.834		
3	12.85	188.2	1.2	20	8.834		
4	16.59	241	1.5	20	8.834		
5	33.78	489.5	3	20	8.834		
6	71.07	1005	6	20	8.834		
7	89.47	1267	7.5	20	8.834		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-060921.mtw
 Equation: $Q = 0.0526792 \cdot A + 1.89676$
 RSD: 0.929 %
 Correlation coefficient: 0.999953



K3 = 0 K2 = 0 K1 = 0.0526792 K0 = 1.89676

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	11.09	207.8	0.6	20	10.64		
3	21.78	413.9	1.2	20	10.64		
4	27.65	532.5	1.5	20	10.64		
5	55.06	1088	3	20	10.64		
6	108.7	2242	6	20	10.64		
7	130.5	2817	7.5	20	10.64		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-060921.mtw
 Equation: $Q = 0.260562 \cdot A + 6.03586$
 RSD: 1.294 %
 Correlation coefficient: 0.999909



K3 = 0 K2 = 0 K1 = 0.260562 K0 = 6.03586

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.138	142.7	2	20			14
3	14.34	281	4	20			14
4	18.67	361.8	5	20			14
5	38.63	729.9	10	20			14
6	81.71	1507	20	20			14
7	103.7	1905	25	20			14

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-060921.mtw
 Equation: $Q = 0.045484 \cdot A + 1.87979$
 RSD: 1.594 %
 Correlation coefficient: 0.999861



K3 = 0 K2 = 0 K1 = 0.045484 K0 = 1.87979

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	8.695	200.9	0.5	20	14.79		
3	17.38	394.8	1	20	14.79		
4	22.59	508	1.25	20	14.79		
5	46.56	1033	2.5	20	14.79		
6	98.65	2147	5	20	14.79		
7	125.6	2723	6.25	20	14.79		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-060921.mtw
 Equation: $Q = 0.0932001 \cdot A - 8.12699$
 RSD: 1.344 %
 Correlation coefficient: 0.999902



K3 = 0 K2 = 0 K1 = 0.0932001 K0 = -8.12699

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.496	280.3	1	20	21.2		
3	14.27	514.4	2	20	21.2		
4	18.27	644.1	2.5	20	21.2		
5	33.95	1169	5	20	21.2		
6	67.72	2234	10	20	21.2		
7	85.25	2763	12.5	20	21.2		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-060921.mtw
 Equation: $Q = 0.148453 \cdot A + 12.5141$
 RSD: 1.833 %
 Correlation coefficient: 0.999817



K3 = 0 K2 = 0 K1 = 0.148453 K0 = 12.5141

Base: Area

Ref.channel: ch1

ISTD:

Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	12.03	368.6	3	20	24.73		
3	23.72	718.8	6	20	24.73		
4	30.61	921.6	7.5	20	24.73		
5	63.34	1883	15	20	24.73		
6	133.7	3939	30	20	24.73		
7	169.9	5000	37.5	20	24.73		

Report date: 6/15/2021 9:08:03 AM
Printed by: wet


Ident: STD1
Analysis from: 6/9/2021 10:00:37 AM
File: _2021-06-09_

Last save: 6/15/2021 8:38:25 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27241

Last save: 6/15/2021 8:38:

SAMPLE:
: AM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 6/15/2021 9:08:11 AM
Printed by: wet


Ident: STD2
Analysis from: 6/9/2021 10:31:12 AM
File: _2021-06-09_

Last save: 6/15/2021 8:38:48 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27242

Last save: 6/15/2021 8:38:

SAMPLE:
: AM/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.73	0.403	2.88	5.16	91.853	6.62
2	8.70	0.208	6.48	11.61	96.044	6.92
3	10.48	0.255	11.09	19.86	207.763	14.97
4	13.43	0.290	7.14	12.79	142.715	10.28
5	15.46	0.335	8.69	15.58	200.922	14.47
6	21.27	0.482	7.50	13.43	280.346	20.19
7	25.12	0.459	12.03	21.56	368.593	26.55
7	28.00	0.348	55.81	99.99	1388.236	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/15/2021 9:08:21 AM
Printed by: wet


Ident: STD3
Analysis from: 6/9/2021 11:01:47 AM
File: _2021-06-09_

Last save: 6/15/2021 8:39:08 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27243

Last save: 6/15/2021 8:39:

SAMPLE:
: AM/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.71	0.362	6.11	5.53	185.677	6.89
2	8.66	0.207	12.84	11.63	188.177	6.98
3	10.42	0.259	21.78	19.72	413.864	15.35
4	13.34	0.284	14.34	12.99	281.034	10.42
5	15.33	0.330	17.37	15.73	394.752	14.64
6	21.24	0.467	14.27	12.92	514.355	19.07
7	25.09	0.456	23.72	21.47	718.847	26.66
7	28.00	0.338	110.44	99.99	2696.706	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/15/2021 9:08:28 AM
Printed by: wet


Ident: STD4
Analysis from: 6/9/2021 11:32:23 AM
File: _2021-06-09_

Last save: 6/15/2021 8:39:21 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27244

Last save: 6/15/2021 8:39:

SAMPLE:
: AM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.71	0.367	7.87	5.53	237.033	6.88
2	8.65	0.206	16.59	11.66	240.992	6.99
3	10.39	0.264	27.65	19.44	532.494	15.45
4	13.28	0.281	18.67	13.12	361.830	10.50
5	15.25	0.327	22.59	15.88	508.031	14.74
6	21.24	0.460	18.27	12.84	644.148	18.69
7	25.09	0.453	30.61	21.52	921.623	26.74
7	28.00	0.337	142.27	100.00	3446.151	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/15/2021 9:08:34 AM
Printed by: wet


Ident: STD5
Analysis from: 6/9/2021 12:02:58 PM
File: _2021-06-09_

Last save: 6/15/2021 8:39:36 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27245

Last save: 6/15/2021 8:39:

SAMPLE:
: AM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.73	0.380	17.13	5.94	506.765	7.34
2	8.64	0.205	33.78	11.71	489.536	7.09
3	10.36	0.271	55.06	19.09	1087.943	15.77
4	13.25	0.275	38.63	13.39	729.919	10.58
5	15.18	0.322	46.56	16.14	1033.376	14.98
6	21.25	0.452	33.95	11.77	1168.842	16.94
7	25.11	0.447	63.34	21.96	1883.474	27.30
7	28.00	0.336	288.45	100.00	6899.855	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/15/2021 9:08:41 AM
Printed by: wet


Ident: STD6
Analysis from: 6/9/2021 12:33:34 PM
File: _2021-06-09_

Last save: 6/15/2021 8:40:04 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27246

Last save: 6/15/2021 8:40:

SAMPLE:
: AM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.72	0.343	36.83	6.15	999.526	7.10
2	8.62	0.199	71.07	11.88	1005.403	7.14
3	10.30	0.284	108.74	18.17	2242.429	15.93
4	13.19	0.267	81.71	13.65	1506.852	10.71
5	15.08	0.313	98.65	16.48	2147.101	15.26
6	21.22	0.437	67.72	11.32	2234.236	15.87
7	25.07	0.441	133.69	22.34	3938.831	27.99
7	28.00	0.326	598.39	100.00	14074.378	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/15/2021 9:08:49 AM
Printed by: wet


Ident: STD7
Analysis from: 6/9/2021 1:04:09 PM
File: _2021-06-09_

Last save: 6/15/2021 8:40:19 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27247

Last save: 6/15/2021 8:40:

SAMPLE:
: AM/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.72	0.360	44.32	5.92	1246.720	7.04
2	8.61	0.199	89.47	11.95	1267.072	7.15
3	10.29	0.301	130.45	17.43	2816.677	15.89
4	13.19	0.266	103.66	13.85	1904.633	10.75
5	15.07	0.312	125.61	16.78	2723.150	15.37
6	21.21	0.433	85.25	11.39	2762.803	15.59
7	25.06	0.441	169.86	22.69	5000.380	28.22
7	28.00	0.330	748.61	100.00	17721.435	100.00

This report has been created by IC Net
METROHM LTD

Date - 6/9/2021

Instrument ID

IC-2

Analyst -

AM

Anion Calibration std - relative error calculations, acceptance limit is -30 to 30 for std 2 and -10 to 10 for other std																		
Parameter	TV of std2	Result of std2	% RE std2	TV of std3	Result of std3	% RE std3	TV of std4	Result of std4	% RE std4	TV of std5	Result of std5	% RE std5	TV of std6	Result of std6	% RE std6	TV of std7	Result of std7	% RE std7
F	0.4	0.405	1.3	0.8	0.777	-2.9	1	0.981	-1.9	2	2.050	2.5	4	4.004	0.1	5	4.984	-0.3
CL	0.6	0.646	7.7	1.2	1.187	-1.1	1.5	1.497	-0.2	3	2.957	-1.4	6	5.988	-0.2	7.5	7.525	0.3
NO2	0.6	0.642	7.0	1.2	1.185	-1.2	1.5	1.497	-0.2	3	2.960	-1.3	6	6.001	0.0	7.5	7.514	0.2
BR	2	2.161	8.1	4	3.963	-0.9	5	5.016	0.3	10	9.811	-1.9	20	19.933	-0.3	25	25.116	0.5
NO3	0.5	0.551	10.2	1	0.992	-0.8	1.25	1.249	-0.1	2.5	2.444	-2.2	5	4.977	-0.5	6.25	6.287	0.6
HPO4	1	0.900	-10.0	2	1.991	-0.4	2.5	2.595	3.8	5	5.040	0.8	10	10.005	0.1	12.5	12.468	-0.3
SO4	3	3.362	12.1	6	5.961	-0.6	7.5	7.467	-0.4	15	14.606	-2.6	30	29.862	-0.5	37.5	37.742	0.6

06/09/2021
AM

Report date: 6/15/2021 9:10:14 AM
Printed by: wet


Ident: ICV
Analysis from: 6/9/2021 1:34:45 PM
File: _2021-06-09_

Last save: 6/15/2021 8:40:33 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27248

Last save: 6/15/2021 8:40:

SAMPLE:
: AM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.74	0.383	16.84	6.00	501.610	7.46
2	8.64	0.205	32.68	11.64	474.064	7.05
3	10.35	0.268	53.21	18.95	1053.709	15.66
4	13.24	0.274	37.59	13.39	706.320	10.50
5	15.17	0.321	45.30	16.13	1001.043	14.88
6	21.27	0.452	33.71	12.00	1166.477	17.34
7	25.12	0.447	61.45	21.88	1825.061	27.13
7	28.00	0.336	280.78	100.00	6728.285	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/15/2021 9:10:23 AM
Printed by: wet


Ident: ICB
Analysis from: 6/9/2021 2:05:21 PM
File: _2021-06-09_

Last save: 6/15/2021 8:41:08 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27249

Last save: 6/15/2021 8:41:

SAMPLE:
: AM/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:14:11 PM
Printed by: wet


Ident: CCV
Analysis from: 6/24/2021 10:07:47 AM
File: _2021-06-24_

Last save: 6/24/2021 10:35:47 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27351

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.83	0.415	14.99	5.71	510.364	7.59
2	8.93	0.209	32.28	12.29	488.418	7.27
3	10.80	0.277	50.22	19.12	1045.246	15.55
4	14.01	0.303	34.11	12.99	726.353	10.81
5	16.15	0.361	40.36	15.37	1015.692	15.11
6	21.37	0.461	30.51	11.62	1067.348	15.88
7	25.24	0.460	60.15	22.90	1868.682	27.80
7	28.00	0.355	262.62	100.00	6722.103	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:14:18 PM
Printed by: wet


Ident: CCB
Analysis from: 6/24/2021 10:43:20 AM
File: _2021-06-24_

Last save: 6/24/2021 11:11:21 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27352

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:14:26 PM
Printed by: wet


Ident: LB115097BLW
Analysis from: 6/24/2021 11:13:59 AM
File: _2021-06-24_

Last save: 6/25/2021 10:34:41 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27353

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 13
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:14:34 PM
Printed by: wet


Ident: LB115097BSW
Analysis from: 6/24/2021 11:44:37 AM
File: _2021-06-24_

Last save: 6/25/2021 10:34:41 AM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27354

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.409	15.13	5.68	508.006	7.53
2	8.90	0.208	32.47	12.20	487.462	7.23
3	10.75	0.276	50.61	19.02	1045.252	15.50
4	13.92	0.298	34.70	13.04	725.639	10.76
5	16.04	0.354	41.15	15.47	1019.777	15.12
6	21.37	0.458	31.47	11.83	1090.662	16.17
7	25.23	0.457	60.57	22.76	1868.186	27.70
7	28.00	0.352	266.08	100.00	6744.985	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:15:06 PM
Printed by: wet


Ident: M2846-03
Analysis from: 6/24/2021 5:14:41 PM
File: _2021-06-24_

Last save: 6/24/2021 5:42:41 PM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27358

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.86	0.549	0.55	0.10	19.336	0.22
2	8.90	0.224	516.43	97.15	8379.927	94.84
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.22	0.351	3.44	0.65	83.596	0.95
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.29	0.469	11.13	2.09	353.154	4.00
7	28.00	0.228	531.56	100.00	8836.013	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:15:49 PM
Printed by: wet


Ident: M2846-03MS
Analysis from: 6/24/2021 5:45:19 PM
File: _2021-06-24_

Last save: 6/24/2021 6:13:19 PM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27359

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.481	16.52	2.07	621.329	3.96
2	8.89	0.219	548.86	68.93	8668.196	55.29
3	10.74	0.254	25.47	3.20	417.874	2.67
4	13.99	0.287	36.51	4.59	732.919	4.67
5	16.09	0.341	45.76	5.75	1088.981	6.95
6	21.53	0.508	41.76	5.24	1763.176	11.25
7	25.38	0.452	72.80	9.14	2218.721	14.15
7	28.00	0.363	787.68	98.92	15511.196	98.94

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:15:57 PM
Printed by: wet


Ident: M2846-03MSD
Analysis from: 6/24/2021 6:15:57 PM
File: _2021-06-24_

Last save: 6/24/2021 6:43:57 PM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27360

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.430	17.81	2.17	622.753	3.88
2	8.87	0.217	557.61	67.91	8710.687	54.22
3	10.71	0.252	27.16	3.31	442.410	2.75
4	13.87	0.277	38.81	4.73	749.404	4.66
5	15.93	0.338	48.40	5.89	1141.999	7.11
6	21.41	0.505	44.65	5.44	1871.555	11.65
7	25.25	0.449	76.41	9.31	2313.026	14.40
7	28.00	0.353	810.86	98.76	15851.835	98.66

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:16:24 PM
Printed by: wet


Ident: CCV
Analysis from: 6/24/2021 7:47:54 PM
File: _2021-06-24_

Last save: 6/25/2021 4:16:24 PM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27363

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.379	16.27	5.85	508.602	7.33
2	8.85	0.204	33.55	12.06	492.579	7.10
3	10.66	0.269	52.57	18.90	1054.834	15.20
4	13.77	0.290	35.96	12.93	727.612	10.48
5	15.84	0.344	42.92	15.43	1029.474	14.83
6	21.36	0.463	35.00	12.58	1249.835	18.01
7	25.23	0.451	61.90	22.25	1878.175	27.06
7	28.00	0.343	278.16	100.00	6941.112	100.00

This report has been created by IC Net
METROHM LTD

Report date: 6/25/2021 4:16:32 PM
Printed by: wet


Ident: CCB
Analysis from: 6/24/2021 8:18:32 PM
File: _2021-06-24_

Last save: 6/24/2021 8:46:32 PM

Method: AnionsIC2-060921.mtw
Run operator: wet
Analysis number: 27364

Last save: 6/17/2021 9:49:

SAMPLE:
: AP/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD