



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115100

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	06/25/2021	12:20
2	M2812-07		1	8	100.00	>212.0	>212.0	06/25/2021	12:45
3	M2812-07DUP		1	8	100.00	>212.0	>212.0	06/25/2021	13:10
4	M2826-02		1	9	100.00	>212.0	>212.0	06/25/2021	13:35
5	M2845-01		1	9	100.00	>212.0	>212.0	06/25/2021	14:05
6	M2845-02		1	8	100.00	>212.0	>212.0	06/25/2021	14:25
7	M2846-03		1	9	100.00	>212.0	>212.0	06/25/2021	14:50
8	M2847-02		1	9	100.00	>212.0	>212.0	06/25/2021	15:10
9	M2848-01		1	8	100.00	>212.0	>212.0	06/25/2021	15:35
10	M2849-01		1	9	100.00	>212.0	>212.0	06/25/2021	15:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 115700

WorkList Name : fp62521 WorkList ID : 150108 Department : Wet-Chemistry Date : 06-25-2021 08:59:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/29/2021	Water	M2812-07	Flash Point	Cool 4 deg C	PSEG03	J12	KG-WATER	06/22/2021	1010A
06/28/2021	Water	M2826-02	Flash Point	Cool 4 deg C	STAN07	J22	LS-1(LIQUID-BLUE)	06/23/2021	1010A
07/01/2021	Water	M2845-01	Flash Point	Cool 4 deg C	PSEG03	J23	41234	06/24/2021	1010A
07/01/2021	Water	M2845-02	Flash Point	Cool 4 deg C	PSEG03	J23	41235	06/24/2021	1010A
07/08/2021	Water	M2846-03	Flash Point	Cool 4 deg C	PSEG03	J22	DRUM-1	06/24/2021	1010A
06/29/2021	Water	M2847-02	Flash Point	Cool 4 deg C	PSEG03	J23	1928-29	06/24/2021	1010A
07/01/2021	Water	M2848-01	Flash Point	Cool 4 deg C	PSEG03	J23	OILY-WATER-DRUM	06/24/2021	1010A
06/29/2021	Water	M2849-01	Flash Point	Cool 4 deg C	PSEG03	J23	NWB-1613	06/24/2021	1010A

Date/Time 06/25/21 12:00
Raw Sample Received by: FS (Webster)
Raw Sample Relinquished by: JACOB ORSON

Date/Time 06/25/21 16:00
Raw Sample Received by: JACOB ORSON
Raw Sample Relinquished by: FS (Webster)