



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115116

Slope : 98.1

BalanceID: WC SC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/26/2021	11:15
2	CAL2	1	Water	NA	NA	20.3	7.00	06/26/2021	11:16
3	CAL3	1	Water	NA	NA	20.3	10.02	06/26/2021	11:19
4	ICV	1	Water	NA	NA	20.4	7.00	06/26/2021	11:20
5	CCV1	1	Water	NA	NA	20.4	2.01	06/26/2021	11:22
6	M2868-01	1	Solid	20.02	20	22.7	6.35	06/26/2021	11:25
7	M2868-01DUP	1	Solid	20.03	20	22.8	6.36	06/26/2021	11:26
8	CCV2	1	Water	NA	NA	20.3	12.02	06/26/2021	11:28

115116

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2368 WorkList ID : 150149 Department : Wet-Chemistry Date : 06-26-2021 11:05:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/25/2021	Solid	M2868-01	pH	Cool 4 deg C	PSEG03	J31	OILY-DEBRIS-TRE-1169	06/25/2021	9045D

Date/Time 06/26/21 11:10
 Raw Sample Received by: JB GWT (CRM)
 Raw Sample Relinquished by: RCCEAD LAB

Date/Time 06/26/21 12:30
 Raw Sample Received by: RCCEAD LAB
 Raw Sample Relinquished by: JB GWT (CRM)