



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115117

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/26/2021	11:15
2	CAL2	1	Water	NA	NA	20.3	7.00	06/26/2021	11:16
3	CAL3	1	Water	NA	NA	20.3	10.03	06/26/2021	11:19
4	ICV	1	Water	NA	NA	20.4	7.00	06/26/2021	11:20
5	CCV1	1	Water	NA	NA	20.4	2.01	06/26/2021	11:22
6	M2868-03	1	Water	NA	NA	20.6	7.18	06/26/2021	11:30
7	M2868-03DUP	1	Water	NA	NA	20.7	7.19	06/26/2021	11:31
8	CCV2	1	Water	NA	NA	20.6	12.02	06/26/2021	11:32

VB 115117

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2368 w WorkList ID : 150150 Department : Wet-Chemistry Date : 06-26-2021 11:06:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/25/2021	Water	M2868-03	pH	Cool 4 deg C	PSEG03	J31	WASTE-WATER-TRE-12C	06/25/2021	9040C

Date/Time 06/26/21 11:10
Raw Sample Received by: JD Wt (200)
Raw Sample Relinquished by: RC (EXT) (56)

Date/Time 06/26/21 12:30
Raw Sample Received by: RC (EXT) (56)
Raw Sample Relinquished by: JD Wt (179)