



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115118

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/26/2021	11:15
2	CAL2	1	Water	NA	NA	20.3	7.00	06/26/2021	11:16
3	CAL3	1	Water	NA	NA	20.3	10.02	06/26/2021	11:19
4	ICV	1	Water	NA	NA	20.4	7.00	06/26/2021	11:20
5	CCV1	1	Water	NA	NA	20.4	2.01	06/26/2021	11:22
6	M2846-02	1	Solid	20.02	20	21.7	7.58	06/26/2021	11:40
7	M2846-02DUP	1	Solid	20.03	20	21.8	7.59	06/26/2021	11:41
8	M2856-21	1	Solid	20.04	20	21.8	6.74	06/26/2021	11:44
9	M2856-22	1	Solid	20.03	20	21.5	6.90	06/26/2021	11:45
10	M2856-23	1	Solid	20.04	20	21.4	6.48	06/26/2021	11:47
11	M2856-24	1	Solid	20.03	20	21.0	6.87	06/26/2021	11:48
12	M2856-25	1	Solid	20.03	20	21.4	7.80	06/26/2021	11:50
13	M2856-26	1	Solid	20.04	20	21.4	8.15	06/26/2021	12:00
14	M2856-27	1	Solid	20.02	20	21.1	7.48	06/26/2021	12:01
15	M2856-28	1	Solid	20.03	20	20.9	5.45	06/26/2021	12:02
16	CCV2	1	Water	NA	NA	20.4	12.02	06/26/2021	12:05
17	M2856-29	1	Solid	20.04	20	20.9	5.45	06/26/2021	12:10
18	M2856-30	1	Solid	20.02	20	21.7	6.80	06/26/2021	12:11
19	M2857-17	1	Solid	20.03	20	20.7	6.73	06/26/2021	12:15
20	M2857-18	1	Solid	20.04	20	21.8	7.14	06/26/2021	12:19

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M2857-19	1	Solid	20.02	20	21.6	7.24	06/26/2021	12:20
22	M2857-20	1	Solid	20.03	20	21.7	7.62	06/26/2021	12:22
23	M2857-21	1	Solid	20.04	20	21.0	10.94	06/26/2021	12:24
24	M2857-22	1	Solid	20.03	20	21.9	9.06	06/26/2021	12:35
25	M2857-23	1	Solid	20.04	20	21.1	8.17	06/26/2021	12:36
26	M2857-24	1	Solid	20.02	20	21.6	8.15	06/26/2021	12:39
27	CCV3	1	Water	NA	NA	20.4	2.01	06/26/2021	12:40
28	M2857-24DUP	1	Solid	20.03	20	21.7	8.16	06/26/2021	11:41
29	CCV4	1	Water	NA	NA	20.6	12.02	06/26/2021	11:42

VB 115118

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2856 WorkList ID : 150151 Department : Wet-Chemistry Date : 06-26-2021 11:07:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/24/2021	Solid	M2846-02	Corrosivity	Cool 4 deg C	PSEG03	J22	1-4-FEET-TP	06/24/2021	9045D
06/21/2021	Solid	M2856-21	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-01-COMP	06/21/2021	9045D
06/21/2021	Solid	M2856-22	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-02-COMP	06/21/2021	9045D
06/21/2021	Solid	M2856-23	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-03-COMP	06/21/2021	9045D
06/21/2021	Solid	M2856-24	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-04-COMP	06/21/2021	9045D
06/22/2021	Solid	M2856-25	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-05-COMP	06/22/2021	9045D
06/22/2021	Solid	M2856-26	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-06-COMP	06/22/2021	9045D
06/22/2021	Solid	M2856-27	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-07-COMP	06/22/2021	9045D
06/22/2021	Solid	M2856-28	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-08-COMP	06/22/2021	9045D
06/22/2021	Solid	M2856-29	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-09-COMP	06/22/2021	9045D
06/22/2021	Solid	M2856-30	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-10-COMP	06/22/2021	9045D
06/22/2021	Solid	M2857-17	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-11-COMP	06/22/2021	9045D
06/22/2021	Solid	M2857-18	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-12-COMP	06/22/2021	9045D
06/22/2021	Solid	M2857-19	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-13-COMP	06/22/2021	9045D
06/22/2021	Solid	M2857-20	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-14-COMP	06/22/2021	9045D
06/17/2021	Solid	M2857-21	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-15-COMP	06/17/2021	9045D
06/23/2021	Solid	M2857-22	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-16-COMP	06/23/2021	9045D
06/23/2021	Solid	M2857-23	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-17-COMP	06/23/2021	9045D
06/23/2021	Solid	M2857-24	Corrosivity	Cool 4 deg C	LIRO01	J41	SB-18-COMP	06/23/2021	9045D

Date/Time 06/26/21 11:10
 Raw Sample Received by: JG Wt Henry
 Raw Sample Relinquished by: RECET 56

Date/Time 06/26/21 12:40
 Raw Sample Received by: PC CEX + 56
 Raw Sample Relinquished by: JG Wt Henry