



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115137

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	06/28/2021	15:15
2	M2868-03		1	9	100.00	>212.0	>212.0	06/28/2021	15:40
3	M2868-03DUP		1	9	100.00	>212.0	>212.0	06/28/2021	16:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB115137

WorkList Name : FP62521-1 WorkList ID : 150135 Department : Wet-Chemistry Date : 06-25-2021 16:25:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/30/2021	Water	M2868-03	Flash Point	Cool 4 deg C	PSEG03	J31	WASTE-WATER-TRE-12C	06/25/2021	1010A

Date/Time 06/28/21 14:50
 Raw Sample Received by: KS (Weber)
 Raw Sample Relinquished by: KS (Weber)

Date/Time 06/28/21 16:10
 Raw Sample Received by: KS (Weber)
 Raw Sample Relinquished by: KS (Weber)