

Analytical Summary ReportAnalysis Method: SM5220 DANALYST: apatelParameter: ASTM CODSUPERVISOR REVIEW BY: ketankumarRun Number: LB115140

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP90886
COD calibration std. 100 ppm	WP90884
COD calibration std. 10 ppm	WP90882
COD calibration std. 150 ppm	WP90885
COD calibration std. 50 ppm	WP90883
COD calibration std. 0 ppm	WP90881
COD CCV std, 50ppm	WP90888
COD Digestion Vials Low Level 0-150Mg/L	W2782
COD ICV-LCS std, 50ppm	WP91808
COD CCV std, 50ppm	WP91807
COD Digestion Vials Low Level 0-150Mg/L	W2824

Temp In (C): <u>150</u>	Date In: <u>06/28/2021</u>	Time In: <u>10:05</u>
Temp Out (C): <u>150</u>	Date Out: <u>06/28/2021</u>	Time Out: <u>12:06</u>

Intercept: 0.2708Slope: 1.015Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.267		05/08/2021	12:05
2	CAL2	10	1	Water	11.000	10.571	-5.7	05/08/2021	12:05
3	CAL3	50	1	Water	51.000	49.98	0	05/08/2021	12:06
4	CAL4	100	1	Water	101.000	99.241	0.8	05/08/2021	12:06
5	CAL5	150	1	Water	153.000	150.472	-0.3	05/08/2021	12:07



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: apatel

Parameter: ASTM COD

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115140

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	49.980	06/28/2021	12:33
2	ICB		NA	NA	1	Water	0.000	-0.267	06/28/2021	12:33
3	CCV1	50	NA	NA	1	Water	50.000	48.994	06/28/2021	12:40
4	CCB1		NA	NA	1	Water	0.000	-0.267	06/28/2021	12:40
5	LB115140BL		NA	NA	1	Water	1.000	0.718	06/28/2021	12:41
6	LB115140BS	50	NA	NA	1	Water	49.000	48.009	06/28/2021	12:41
7	M2860-01		NA	NA	1	Water	7.000	6.630	06/28/2021	12:42
8	M2860-01DUP		NA	NA	1	Water	6.000	5.645	06/28/2021	12:42
9	M2860-01MS	50	NA	NA	1	Water	51.000	49.980	06/28/2021	12:45
10	M2860-01MSD	50	NA	NA	1	Water	52.000	50.965	06/28/2021	12:45
11	CCV2	50	NA	NA	1	Water	49.000	48.009	06/28/2021	12:46
12	CCB2		NA	NA	1	Water	0.000	-0.267	06/28/2021	12:46