CHEMTECH

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB115142

Analyst By : jignesh
Supervisor Review By : apatel
Slope : 99.1
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	06/29/2021	08:15
2	CAL2	1	Water	NA	NA	21.2	7.01	06/29/2021	08:16
3	CAL3	1	Water	NA	NA	21.2	10.03	06/29/2021	08:17
4	ICV	1	Water	NA	NA	21.4	6.99	06/29/2021	08:19
5	CCV1	1	Water	NA	NA	21.2	2.01	06/29/2021	08:20
6	M2882-03	1	Water	NA	NA	24.6	4.92	06/29/2021	08:22
7	M2882-05	1	Water	NA	NA	24.7	7.84	06/29/2021	08:24
8	M2882-07	1	Water	NA	NA	24.7	5.65	06/29/2021	08:26
9	M2883-01	1	Water	NA	NA	24.7	12.47	06/29/2021	08:30
10	M2883-01DUP	1	Water	NA	NA	24.8	12.48	06/29/2021	08:31
11	CCV2	1	Water	NA	NA	21.4	12.02	06/29/2021	08:33

28 115142

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2882 w **WorkList ID :** 150186 **Department :** Wet-Chemistry **Date :** 06-29-2021 08:04:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/28/2021	Water	M2882-03 C	pH	Cool 4 deg C	PSEG03	J32	DRIP-COMP-AUD-1121-1	06/28/2021	9040C
06/28/2021	Water	M2882-05 C	pH	Cool 4 deg C	PSEG03	J32	SEALPOT-AUD-1123	06/28/2021	9040C
06/28/2021	Water	M2882-07 C	pH	Cool 4 deg C	PSEG03	J32	AUD-1126	06/28/2021	9040C
06/28/2021	Water	M2883-01 C	pH	Cool 4 deg C	PSEG03	J33	GLYCOL	06/28/2021	9040C

Date/Time 06/29/21 08:05

Raw Sample Received by: JP (wet chem)

Raw Sample Relinquished by: SH

Date/Time 06/29/21 10:00

Raw Sample Received by: SH

Raw Sample Relinquished by: JP (wet chem)