



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115143

Slope : 99.1

BalanceID:

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	06/29/2021	08:15
2	CAL2	1	Water	NA	NA	21.2	7.01	06/29/2021	08:16
3	CAL3	1	Water	NA	NA	21.2	10.03	06/29/2021	08:17
4	ICV	1	Water	NA	NA	21.4	6.99	06/29/2021	08:19
5	CCV1	1	Water	NA	NA	21.2	2.01	06/29/2021	08:20
6	M2882-01	1	Solid	20.02	20	24.3	7.49	06/29/2021	08:37
7	M2882-01DUP	1	Solid	20.03	20	24.4	7.50	06/29/2021	08:38
8	CCV2	1	Water	NA	NA	21.2	12.02	06/29/2021	08:40

WORKLIST(Hardcopy Internal Chain)

18 115143

WorkList Name : ph m2882 s WorkList ID : 150187 Department : Wet-Chemistry Date : 06-29-2021 08:05:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/28/2021	Solid	M2882-01	pH	Cool 4 deg C	PSEG03	J32	SPILL-COMP-31277-AUC	06/28/2021	9045D

Date/Time 06/29/21 08:05
 Raw Sample Received by: JD cut-chem
 Raw Sample Relinquished by: JD (sus)

Date/Time 06/24/21 10:00
 Raw Sample Received by: JD (sus)
 Raw Sample Relinquished by: JD cut-chem