



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115146

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	06/29/2021	12:20
2	M2882-03		1	9	100.00	>212.0	>212.0	06/29/2021	12:45
3	M2882-03DUP		1	9	100.00	>212.0	>212.0	06/29/2021	13:05
4	M2882-05		1	8	100.00	>212.0	>212.0	06/29/2021	13:35
5	M2882-07		1	9	100.00	>212.0	>212.0	06/29/2021	13:58
6	M2883-01		1	8	100.00	>212.0	>212.0	06/29/2021	14:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp62921 WorkList ID : 150203 Department : Wet-Chemistry Date : 06-29-2021 09:50:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/01/2021	Water	M2882-03	Flash Point	Cool 4 deg C	PSEG03	J32	DRIP-COMP-AUD-1121-1	06/28/2021	1010A
07/01/2021	Water	M2882-05	Flash Point	Cool 4 deg C	PSEG03	J32	SEALPOT-AUD-1123	06/28/2021	1010A
07/01/2021	Water	M2882-07	Flash Point	Cool 4 deg C	PSEG03	J32	AUD-1126	06/28/2021	1010A
07/01/2021	Water	M2883-01	Flash Point	Cool 4 deg C	PSEG03	J33	GLYCOL	06/28/2021	1010A

Date/Time 06/29/21 12:00
 Raw Sample Received by: KS (Welch)
 Raw Sample Relinquished by: JP at (nem)

Date/Time 06/29/21 14:30
 Raw Sample Received by: JP at (nem)
 Raw Sample Relinquished by: KS (Welch)