



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 6/30/2021

OVENTEMP IN Celsius(°C): 108
Time IN: 16:20
In Date: 06/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 06/30/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115150

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2893-01	BULL-WASTE-YELLOW-CAN	1	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
M2894-01	PAT-TP	2	1.14	9.73	9.29	94.9	
M2894-02	PAT-TP	3	1.12	9.69	9.03	92.3	
M2894-03	PAT-TP-EPH-2	4	1.18	9.96	9.29	92.4	
M2894-04	PSCON-1	5	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2894-05	PSCON-2	6	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2895-01	COMP-1	7	1.14	9.77	8.43	84.5	
M2895-02	COMP-2	8	1.13	9.68	8.24	83.2	
M2895-02D	COMP-2DUP	9	1.15	9.7	8.26	83.2	
M2895-03	COMP-3	10	1.18	9.59	7.74	78.0	
M2896-01	FS-01	11	1.12	9.83	5.59	51.3	
M2896-03	FS-02	12	1.00	2.00	2.00	100.0	TAR SAMPLE
M2896-05	FS-03	13	1.16	9.78	4.41	37.7	
M2896-07	FS-04	14	1.18	9.75	6.3	59.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$