


LB115151
INSTRUMENT ID: QV1

7471B



Linear

A= 0.0000e+000

B= 1.7782e-004

C= -6.4037e-003

Rho= 0.9998243

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.025	0.025	178	0.000	178					
0.2	0.200	0.214	0.014	1238	0.0 %	1238					
2.5	2.500	2.528	0.028	14255	0.0 %	14255					
5.0	5.000	4.943	-0.057	27831	0.0 %	27831					
7.5	7.500	7.386	-0.114	41572	0.0 %	41572					
10.0	10.000	10.104	0.104	56856	0.0 %	56856					

LB115151 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 Jun 2021 15:27:09					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	178	-	0.0000	7471B	PPB	29 Jun 2021 15:33:04	S	Std
0.2 - 1	50.2	1238	-	0.2000	7471B	PPB	29 Jun 2021 15:35:20	S	Std
2.5 - 1	52.5	14255	-	2.5000	7471B	PPB	29 Jun 2021 15:37:38	S	Std
5.0 - 1	55	27831	-	5.0000	7471B	PPB	29 Jun 2021 15:39:54	S	Std
7.5 - 1	57.5	41572	-	7.5000	7471B	PPB	29 Jun 2021 15:42:11	S	Std
10.0 - 1	510	56856	-	10.0000	7471B	PPB	29 Jun 2021 15:47:24	S	Std
ICV81 - 1	ICV81	22352	3.9683	-	7471B	PPB	29 Jun 2021 15:50:19	U	SMPL
ICB81 - 1	ICB81	-31	-0.0119	-	7471B	PPB	29 Jun 2021 15:52:34	U	SMPL
CCV85 - 1	CCV85	27746	4.9275	-	7471B	PPB	29 Jun 2021 15:54:50	U	SMPL
CCB85 - 1	CCB85	-21	-0.0101	-	7471B	PPB	29 Jun 2021 15:57:04	U	SMPL
CRA - 1	CRA	1235	0.2132	-	7471B	PPB	29 Jun 2021 15:59:20	U	SMPL
HighStd - 1	HighStd	56789	10.0920	-	7471B	PPB	29 Jun 2021 16:01:35	U	SMPL
ChkStd - 1	ChkStd	38678	6.8714	-	7471B	PPB	29 Jun 2021 16:03:50	U	SMPL
PB137424BL - 1	PBS	-72	-0.0192	-	7471B	PPB	29 Jun 2021 16:06:08	U	SMPL
PB137424BS - 1	LCSS	21615	3.8372	-	7471B	PPB	29 Jun 2021 16:08:25	U	SMPL
M2882-01 - 1	SPILL-COMP-31277-AUD-1077	502	0.0829	-	7471B	PPB	29 Jun 2021 16:10:41	U	SMPL
M2884-01 - 1	OK-03-062821	2084	0.3642	-	7471B	PPB	29 Jun 2021 16:12:57	U	SMPL
M2885-01 - 1	26644	5327	0.9409	-	7471B	PPB	29 Jun 2021 16:15:14	U	SMPL
M2885-03 - 1	41547	10437	1.8495	-	7471B	PPB	29 Jun 2021 16:17:31	U	SMPL
M2886-01 - 1	20426	985	0.1688	-	7471B	PPB	29 Jun 2021 16:19:47	U	SMPL
CCV86 - 1	CCV86	27833	4.9429	-	7471B	PPB	29 Jun 2021 16:22:02	U	SMPL
CCB86 - 1	CCB86	-58	-0.0167	-	7471B	PPB	29 Jun 2021 16:24:17	U	SMPL
M2886-03 - 1	26878	4298	0.7579	-	7471B	PPB	29 Jun 2021 16:26:35	U	SMPL
M2886-05 - 1	RBR-24	89310	15.8750	-	7471B	PPB	29 Jun 2021 16:28:50	U	SMPL
M2889-01 - 1	SB-1(11.5-12)	1216	0.2098	-	7471B	PPB	29 Jun 2021 16:31:05	U	SMPL
M2889-02 - 1	SB-2(11.5-12)	572	0.0953	-	7471B	PPB	29 Jun 2021 16:33:23	U	SMPL
M2889-03 - 1	SB-3(11.5-12)	1577	0.2740	-	7471B	PPB	29 Jun 2021 16:35:39	U	SMPL
M2889-04 - 1	SB-4(14-14.5)	431	0.0702	-	7471B	PPB	29 Jun 2021 16:37:56	U	SMPL
M2889-05 - 1	SB-5(17.5-18)	525	0.0870	-	7471B	PPB	29 Jun 2021 16:40:12	U	SMPL
M2889-05DUP - 1	SB-5(17.5-18)DUP	590	0.0985	-	7471B	PPB	29 Jun 2021 16:42:29	U	SMPL
M2889-05MS - 1	SB-5(17.5-18)MS	20066	3.5618	-	7471B	PPB	29 Jun 2021 16:44:46	U	SMPL
M2889-05MSD - 1	SB-5(17.5-18)MSD	21341	3.7885	-	7471B	PPB	29 Jun 2021 16:47:01	U	SMPL
CCV87 - 1	CCV87	27762	4.9303	-	7471B	PPB	29 Jun 2021 16:49:17	U	SMPL
CCB87 - 1	CCB87	-64	-0.0178	-	7471B	PPB	29 Jun 2021 16:51:32	U	SMPL
M2891-01 - 1	TR-04-062821	2012	0.3514	-	7471B	PPB	29 Jun 2021 16:53:49	U	SMPL
M2889-05LX5 - 1		162	0.0224	-	7471B	PPB	29 Jun 2021 16:56:04	U	SMPL
M2889-05A - 1		21999	3.9055	-	7471B	PPB	29 Jun 2021 16:58:20	U	SMPL
M2886-05DLX5 - 1		20867	3.7042	-	7471B	PPB	29 Jun 2021 17:11:13	U	SMPL
CCV88 - 1	CCV88	27970	4.9673	-	7471B	PPB	29 Jun 2021 17:13:28	U	SMPL
CCB88 - 1	CCB88	-21	-0.0101	-	7471B	PPB	29 Jun 2021 17:15:44	U	SMPL