



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115161

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	06/30/2021	10:50
2	CAL2	1	Water	NA	NA	21.4	7.00	06/30/2021	10:51
3	CAL3	1	Water	NA	NA	21.4	10.03	06/30/2021	10:52
4	ICV	1	Water	NA	NA	21.4	6.99	06/30/2021	11:00
5	CCV1	1	Water	NA	NA	21.5	2.01	06/30/2021	11:01
6	M2903-01	1	Water	NA	NA	26.0	5.00	06/30/2021	11:02
7	M2903-01DUP	1	Water	NA	NA	26.1	5.01	06/30/2021	11:03
8	CCV2	1	Water	NA	NA	21.5	12.03	06/30/2021	11:05

115161

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2903 WorkList ID : 150231 Department : Wet-Chemistry Date : 06-30-2021 10:43:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/21/2021	Water	M2903-01	pH	Cool 4 deg C	STER14	J41	JS10000	06/21/2021	9040C

Date/Time 06/30/21 10:43
Raw Sample Received by: JD G-A-Gen
Raw Sample Relinquished by: J.C (sm)

Date/Time 06/30/21 12:30
Raw Sample Received by: J.C (sm)
Raw Sample Relinquished by: JD Out-Open