

LB115167


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46536



A= 0.0000e+000

B= 2.1475e-004

C= -4.4367e-002

Rho= 0.9999329

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.009	0.009	249	0.000	249					-7
0.05	0.050					207					-2
0.20	0.200	0.186	-0.014	1074	0.0 %	1074					1
2.50	2.500	2.451	-0.049	11620	0.0 %	11620					1
5.00	5.000	5.064	0.064	23789	0.0 %	23789					-1
7.50	7.500	7.540	0.040	35316	0.0 %	35316					
10.0	10.000	9.949	-0.051	46536	0.0 %	46536					

LB115167
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	249	-		0 SFAM01.1	PPB	6/30/2021 14:32	Std
	0.2 S01	1074	-		0.2 SFAM01.1	PPB	6/30/2021 14:34	Std
	2.5 S02	11620	-		2.5 SFAM01.1	PPB	6/30/2021 14:36	Std
	5 S03	23789	-		5 SFAM01.1	PPB	6/30/2021 14:39	Std
	7.5 S04	35316	-		7.5 SFAM01.1	PPB	6/30/2021 14:41	Std
	10 S05	46536	-		10 SFAM01.1	PPB	6/30/2021 14:43	Std
ICV049	ICV049	18981	4.0318 -		SFAM01.1	PPB	6/30/2021 14:46	SMPL
ICB049	ICB049	136	-0.0152 -		SFAM01.1	PPB	6/30/2021 14:48	SMPL
CCV093	CCV093	23839	5.0751 -		SFAM01.1	PPB	6/30/2021 14:50	SMPL
CCB093	CCB093	143	-0.0137 -		SFAM01.1	PPB	6/30/2021 14:53	SMPL
PB137442BL	PBS442	210	0.0007 -		SFAM01.1	PPB	6/30/2021 14:55	SMPL
M2785-01	JXMD0	832	0.1343 -		SFAM01.1	PPB	6/30/2021 15:02	SMPL
M2785-02	JXMD0D	899	0.1487 -		SFAM01.1	PPB	6/30/2021 15:04	SMPL
M2785-03	JXMD0S	12275	2.5917 -		SFAM01.1	PPB	6/30/2021 15:06	SMPL
CCV094	CCV094	23546	5.0122 -		SFAM01.1	PPB	6/30/2021 15:09	SMPL
CCB094	CCB094	137	-0.0149 -		SFAM01.1	PPB	6/30/2021 15:11	SMPL