



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115169

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.00	06/30/2021	15:00
2	CAL2	1	Water	NA	NA	21.4	7.01	06/30/2021	15:01
3	CAL3	1	Water	NA	NA	21.4	10.03	06/30/2021	15:02
4	ICV	1	Water	NA	NA	21.5	6.99	06/30/2021	15:05
5	CCV1	1	Water	NA	NA	21.3	2.01	06/30/2021	15:06
6	M2894-02	1	Solid	20.02	20	24.7	9.10	06/30/2021	15:09
7	M2894-02DUP	1	Solid	20.03	20	24.8	9.11	06/30/2021	15:10
8	M2900-01	1	Solid	20.01	20	26.5	9.81	06/30/2021	15:13
9	CCV2	1	Water	NA	NA	21.5	12.02	06/30/2021	15:15

VB 115169

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m2900 WorkList ID : 150237 Department : Wet-Chemistry Date : 06-30-2021 13:45:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/29/2021	Solid	M2894-02	Corrosivity	Cool 4 deg C	PSEG03	J33	PAT-TP	06/29/2021	9045D
06/30/2021	Solid	M2900-01	Corrosivity	Cool 4 deg C	PSEG03	J41	VB16208	06/30/2021	9045D

Date/Time 06/30/21 13:50

Raw Sample Received by: JO (out-Chem)

Raw Sample Relinquished by: JR (SM)

Date/Time 06/30/21 16:30

Raw Sample Received by: JR (SM)

Raw Sample Relinquished by: JO out (Chem)