

LB 115171


7471B

INSTRUMENT 1D : CV2

Linear

 μ Abs.:

53296



A= 0.0000e+000

B= 1.8890e-004

C= -8.5415e-002

Rho= 0.9999630

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.005	-0.005	428	0.000	428					0/65
0.2	0.200	0.163	-0.037	1316	0.0 %	1316					-19
2.5	2.500	2.524	0.024	13815	0.0 %	13815					1
5.0	5.000	5.057	0.057	27225	0.0 %	27225					0
7.5	7.500	7.478	-0.022	40040	0.0 %	40040					0
10.0	10.000	9.982	-0.018	53296	0.0 %	53296					0

LB115171 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 30 Jun 2021 15:55:38				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units				
0.0 - 1	SC	428	-	0.0000	7471B	PPB		30 Jun 2021 15:57:32	S	Std
0.2 - 1	SC-2	1316	-	0.2000	7471B	PPB		30 Jun 2021 15:59:49	S	Std
2.5 - 1	SC-5	13815	-	2.5000	7471B	PPB		30 Jun 2021 16:02:06	S	Std
5.0 - 1	SC-5	27225	-	5.0000	7471B	PPB		30 Jun 2021 16:04:22	S	Std
7.5 - 1	SC-5	40040	-	7.5000	7471B	PPB		30 Jun 2021 16:06:39	S	Std
10.0 - 1	SC-10	53296	-	10.0000	7471B	PPB		30 Jun 2021 16:11:34	S	Std
ICV84 - 1	ICV84	21508	3.9774	-	7471B	PPB		30 Jun 2021 16:14:24	U	SMPL
ICB84 - 1	ICB84	212	-0.0454	-	7471B	PPB		30 Jun 2021 16:16:38	U	SMPL
CCV95 - 1	CCV95	27666	5.1406	-	7471B	PPB		30 Jun 2021 16:18:53	U	SMPL
CCB95 - 1	CCB95	222	-0.0435	-	7471B	PPB		30 Jun 2021 16:21:07	U	SMPL
CRA - 1	CRA	1309	0.1618	-	7471B	PPB		30 Jun 2021 16:26:03	U	SMPL
HighStd - 1	HighStd	54392	10.1890	-	7471B	PPB		30 Jun 2021 16:28:18	U	SMPL
ChkStd - 1	ChkStd	37369	6.9734	-	7471B	PPB		30 Jun 2021 16:30:33	U	SMPL
PB137443BL - 1	PBS	275	-0.0335	-	7471B	PPB		30 Jun 2021 16:32:50	U	SMPL
PB137443BS - 1	LCSS	19710	3.6377	-	7471B	PPB		30 Jun 2021 16:35:07	U	SMPL
M2894-02 - 1	PAT-TP	23474	4.3487	-	7471B	PPB		30 Jun 2021 16:37:24	U	SMPL
M2894-02DUP - 1	PAT-TPDUP	23082	4.2747	-	7471B	PPB		30 Jun 2021 16:39:40	U	SMPL
M2894-02MS - 1	PAT-TPMS	40524	7.5694	-	7471B	PPB		30 Jun 2021 16:41:57	U	SMPL
M2894-02MSD - 1	PAT-TPMSD	43219	8.0785	-	7471B	PPB		30 Jun 2021 16:46:51	U	SMPL
M2895-01 - 1	COMP-1	2535	0.3934	-	7471B	PPB		30 Jun 2021 16:49:06	U	SMPL
CCV96 - 1	CCV96	26851	4.9866	-	7471B	PPB		30 Jun 2021 16:51:22	U	SMPL
CCB96 - 1	CCB96	179	-0.0516	-	7471B	PPB		30 Jun 2021 16:53:39	U	SMPL
M2895-02 - 1	COMP-2	1997	0.2918	-	7471B	PPB		30 Jun 2021 16:55:55	U	SMPL
M2895-03 - 1	COMP-3	3637	0.6016	-	7471B	PPB		30 Jun 2021 16:58:10	U	SMPL
M2896-01 - 1	FS-01	4105	0.6900	-	7471B	PPB		30 Jun 2021 17:00:26	U	SMPL
M2896-03 - 1	FS-02	2503	0.3874	-	7471B	PPB		30 Jun 2021 17:02:41	U	SMPL
M2896-05 - 1	FS-03	1054	0.1137	-	7471B	PPB		30 Jun 2021 17:04:57	U	SMPL
M2896-07 - 1	FS-04	2431	0.3738	-	7471B	PPB		30 Jun 2021 17:07:13	U	SMPL
M2894-02LX5 - 1		4689	0.8003	-	7471B	PPB		30 Jun 2021 17:09:29	U	SMPL
M2894-02A - 1		42386	7.9211	-	7471B	PPB		30 Jun 2021 17:11:46	U	SMPL
CCV97 - 1	CCV97	26185	4.8608	-	7471B	PPB		30 Jun 2021 17:14:03	U	SMPL
CCB97 - 1	CCB97	175	-0.0524	-	7471B	PPB		30 Jun 2021 17:16:19	U	SMPL