



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115172

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.00	06/30/2021	15:00
2	CAL2	1	Water	NA	NA	21.4	7.01	06/30/2021	15:01
3	CAL3	1	Water	NA	NA	21.4	10.03	06/30/2021	15:02
4	ICV	1	Water	NA	NA	21.5	6.99	06/30/2021	15:05
5	CCV1	1	Water	NA	NA	20.3	2.02	06/30/2021	16:50
6	M2908-01	1	Solid	20	20	25.6	7.98	06/30/2021	16:51
7	M2908-01DUP	1	Solid	20	20	25.6	7.99	06/30/2021	16:52
8	CCV2	1	Water	NA	NA	21.4	12.01	06/30/2021	16:53

WORKLIST(Hardcopy Internal Chain)

LB 11 5772

WorkList Name : ph63021 WorkList ID : 150245 Department : Wet-Chemistry Date : 06-30-2021 16:48:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/29/2021	Solid	M2908-01	pH	Cool 4 deg C	ROYF02	J41	WC-CROPSEY-SOLIDS	06/29/2021	9045D

Date/Time 06/30/21 16:49
 Raw Sample Received by: KS (Welcher)
 Raw Sample Relinquished by: JO Gut (Hem)

Date/Time 06/30/21 17:00
 Raw Sample Received by: JO Gut (Hem)
 Raw Sample Relinquished by: KS (Welcher)