



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115173

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.00	06/30/2021	15:00
2	CAL2	1	Water	NA	NA	21.4	7.01	06/30/2021	15:01
3	CAL3	1	Water	NA	NA	21.4	10.03	06/30/2021	15:02
4	ICV	1	Water	NA	NA	21.5	6.99	06/30/2021	15:05
5	CCV1	1	Water	NA	NA	21.3	2.02	06/30/2021	16:54
6	M2908-02	1	Water	NA	NA	16.4	9.53	06/30/2021	16:55
7	M2908-02DUP	1	Water	NA	NA	16.4	9.54	06/30/2021	16:56
8	CCV2	1	Water	NA	NA	21.3	12.01	06/30/2021	16:57

WORKLIST(Hardcopy Internal Chain)

LB115173

WorkList Name : ph63021-1 WorkList ID : 150246 Department : Wet-Chemistry Date : 06-30-2021 16:49:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/29/2021	Water	M2908-02	pH	Cool 4 deg C	ROYF02	J41	WC-CROPSEY-AQUEOU	06/29/2021	9040C

Date/Time 06/30/21 16:50
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: JA G4 (DPM)

Date/Time 06/30/21 17:00
 Raw Sample Received by: JA G4 (DPM)
 Raw Sample Relinquished by: KS (Wetchem)