



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115179

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	07/01/2021	12:50
2	M2908-02		1	9	100.00	>212.0	>212.0	07/01/2021	13:20
3	M2908-02DUP		1	9	100.00	>212.0	>212.0	07/01/2021	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP7121 WorkList ID : 150267 Department : Wet-Chemistry Date : 07-01-2021 11:09:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/07/2021	Water	M2908-02	Flash Point	Cool 4 deg C	ROYF02	J41	WC-CROPSEY-AQUEOU	06/29/2021	1010A

Date/Time 07/01/21 11:15
 Raw Sample Received by: KS (webchm)
 Raw Sample Relinquished by: BH (Ding lab)

Date/Time 07/01/21 13:50
 Raw Sample Received by: BH (Ding lab)
 Raw Sample Relinquished by: KS (webchm)