

LB115182


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51879



A= 0.0000e+000

B= 1.9365e-004

C= -2.0872e-003

Rho= 0.9999592

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.018	0.018	104	0.000	104					
0.2	0.200	0.204	0.004	1063	0.0 %	1063					
2.5	2.500	2.519	0.019	13020	0.0 %	13020					
5.0	5.000	4.943	-0.057	25538	0.0 %	25538					
7.5	7.500	7.472	-0.028	38595	0.0 %	38595					
10.0	10.000	10.044	0.044	51879	0.0 %	51879					

100

LB115182 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Jul 2021 15:21:41

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	104	-	0.0000	7471B	PPB	01 Jul 2021 15:27:45	S	Std
0.2 - 1	50.2	1063	-	0.2000	7471B	PPB	01 Jul 2021 15:30:02	S	Std
2.5 - 1	52.5	13020	-	2.5000	7471B	PPB	01 Jul 2021 15:32:18	S	Std
5.0 - 1	55	25538	-	5.0000	7471B	PPB	01 Jul 2021 15:34:35	S	Std
7.5 - 1	57.5	38595	-	7.5000	7471B	PPB	01 Jul 2021 15:36:51	S	Std
10.0 - 1	510	51879	-	10.0000	7471B	PPB	01 Jul 2021 15:39:08	S	Std
ICV86 - 1	ICV86	19937	3.8586	-	7471B	PPB	01 Jul 2021 15:41:52	U	SMPL
ICB86 - 1	ICB86	5	-0.0011	-	7471B	PPB	01 Jul 2021 15:44:06	U	SMPL
CCV04 - 1	CCV04	25294	4.8960	-	7471B	PPB	01 Jul 2021 15:46:21	U	SMPL
CCB04 - 1	CCB04	2	-0.0017	-	7471B	PPB	01 Jul 2021 15:48:36	U	SMPL
CRA - 1	CRA	1042	0.1997	-	7471B	PPB	01 Jul 2021 15:50:52	U	SMPL
HighStd - 1	HighStd	51768	10.0226	-	7471B	PPB	01 Jul 2021 15:53:07	U	SMPL
ChkStd - 1	ChkStd	34769	6.7308	-	7471B	PPB	01 Jul 2021 15:58:07	U	SMPL
PB137477BL - 1	PBS	-117	-0.0247	-	7471B	PPB	01 Jul 2021 16:00:22	U	SMPL
M2899-01 - 1	ETGI-352	4302	0.8310	-	7471B	PPB	01 Jul 2021 16:04:54	U	SMPL
M2899-01DUP - 1	ETGI-352DUP	3805	0.7347	-	7471B	PPB	01 Jul 2021 16:08:15	U	SMPL
M2899-01MS - 1	ETGI-352MS	23818	4.6102	-	7471B	PPB	01 Jul 2021 16:13:27	U	SMPL
M2899-01MSD - 1	ETGI-352MSD	22688	4.3914	-	7471B	PPB	01 Jul 2021 16:15:44	U	SMPL
M2900-01 - 1	VB16208	2875	0.5546	-	7471B	PPB	01 Jul 2021 16:18:00	U	SMPL
CCV05 - 1	CCV05	24863	4.8125	-	7471B	PPB	01 Jul 2021 16:20:15	U	SMPL
CCB05 - 1	CCB05	4	-0.0013	-	7471B	PPB	01 Jul 2021 16:22:30	U	SMPL
M2901-01 - 1	124035	10864	2.1017	-	7471B	PPB	01 Jul 2021 16:24:46	U	SMPL
M2901-03 - 1	124033	4162	0.8039	-	7471B	PPB	01 Jul 2021 16:27:01	U	SMPL
M2902-01 - 1	CL-01-063021	6821	1.3188	-	7471B	PPB	01 Jul 2021 16:29:16	U	SMPL
PB137477BS - 1	LCSS	19534	3.7806	-	7471B	PPB	01 Jul 2021 16:36:32	U	SMPL
M2899-01LX5 - 1		817	0.1561	-	7471B	PPB	01 Jul 2021 16:38:47	U	SMPL
M2899-01A - 1		21482	4.1578	-	7471B	PPB	01 Jul 2021 16:41:03	U	SMPL
CCV06 - 1	CCV06	24375	4.7180	-	7471B	PPB	01 Jul 2021 16:43:20	U	SMPL
CCB06 - 1	CCB06	10	-0.0002	-	7471B	PPB	01 Jul 2021 16:45:37	U	SMPL