



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115191

Slope : 98.8

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.8	4.01	07/02/2021	09:35
2	CAL2	1	Water	NA	NA	21.6	7.01	07/02/2021	09:36
3	CAL3	1	Water	NA	NA	21.7	10.02	07/02/2021	09:37
4	ICV	1	Water	NA	NA	21.7	7.01	07/02/2021	09:39
5	CCV1	1	Water	NA	NA	21.8	2.02	07/02/2021	09:40
6	M2913-02	1	Solid	20	20	24.5	8.45	07/02/2021	09:42
7	M2913-02DUP	1	Solid	20	20	24.5	8.46	07/02/2021	11:14
8	M2913-04	1	Solid	20	20	24.4	8.75	07/02/2021	11:15
9	M2913-06	1	Solid	20	20	24.6	10.15	07/02/2021	11:16
10	M2913-08	1	Solid	20	20	24.4	8.72	07/02/2021	11:17
11	M2913-10	1	Solid	20	20	24.5	8.98	07/02/2021	11:18
12	M2913-12	1	Solid	20	20	24.4	10.24	07/02/2021	11:19
13	M2913-14	1	Solid	20	20	24.3	8.74	07/02/2021	11:20
14	M2913-16	1	Solid	20	20	24.3	8.86	07/02/2021	11:21
15	M2913-18	1	Solid	20	20	24.2	8.07	07/02/2021	11:22
16	CCV2	1	Water	NA	NA	21.6	12.01	07/02/2021	11:24

LB 115191

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY7221

WorkList ID : 150296

Department : Wet-Chemistry

Date : 07-02-2021 09:05:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/01/2021	Solid	M2913-02	Corrosivity	Cool 4 deg C	TULL02	J42	D-2500	07/01/2021	9045D
07/01/2021	Solid	M2913-04	Corrosivity	Cool 4 deg C	TULL02	J42	D-2700	07/01/2021	9045D
07/01/2021	Solid	M2913-06	Corrosivity	Cool 4 deg C	TULL02	J42	D-3100	07/01/2021	9045D
07/01/2021	Solid	M2913-08	Corrosivity	Cool 4 deg C	TULL02	J42	D-3400	07/01/2021	9045D
07/01/2021	Solid	M2913-10	Corrosivity	Cool 4 deg C	TULL02	J42	D-3700	07/01/2021	9045D
07/01/2021	Solid	M2913-12	Corrosivity	Cool 4 deg C	TULL02	J42	D-4050	07/01/2021	9045D
07/01/2021	Solid	M2913-14	Corrosivity	Cool 4 deg C	TULL02	J42	D-4300	07/01/2021	9045D
07/01/2021	Solid	M2913-16	Corrosivity	Cool 4 deg C	TULL02	J42	P-D20	07/01/2021	9045D
07/01/2021	Solid	M2913-18	Corrosivity	Cool 4 deg C	TULL02	J42	P-D19	07/01/2021	9045D

Date/Time 07/02/21 09:15
Raw Sample Received by: KS (Wetchem)
Raw Sample Relinquished by: PC CEX+ Lab

Date/Time 07/02/21 11:30
Raw Sample Received by: PC CEX+ Lab
Raw Sample Relinquished by: KS (WC)