

Analytical Summary ReportAnalysis Method: SM2130 BANALYST: apatelParameter: TurbiditySUPERVISOR REVIEW BY: ketankumarRun Number: LB115196

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP91844
Turbidity Calibration std, 20NTU	WP91840
Turbidity Calibration std, 80NTU	WP91838
Turbidity Calibration std, 40NTU	WP91839
Turbidity Calibration std, 1NTU	WP91843
Turbidity Calibration std, 5NTU	WP91842
Turbidity Calibration - CCV std, 10 NTU	WP91841
10 NTU Standard 500 ml	W2752

Intercept: 0.2132Slope: 0.991Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.176	-0.04		07/02/2021	08:45
2	CAL2	1	1	1.190	0.99	1.4	07/02/2021	08:47
3	CAL3	5	1	5.280	5.11	-2.3	07/02/2021	08:49
4	CAL4	10	1	9.950	9.83	1.8	07/02/2021	08:52
5	CAL5	20	1	20.200	20.17	-0.8	07/02/2021	08:54
6	CAL6	40	1	39.800	39.95	0.1	07/02/2021	08:57



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115196

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.178	07/02/2021	09:02
2	ICB		1	0.182	-0.031	07/02/2021	09:05
3	CCV1	10	1	9.920	9.795	07/02/2021	09:08
4	CCB1		1	0.186	-0.027	07/02/2021	09:11
5	LB115196BL		1	0.191	-0.022	07/02/2021	09:14
6	M2920-01		1	39.100	39.240	07/02/2021	09:17
7	M2920-03		1	32.500	32.580	07/02/2021	09:21
8	M2920-03DUP		1	33.100	33.185	07/02/2021	09:24
9	CCV2	10	1	10.100	9.977	07/02/2021	09:27
10	CCB2		1	0.190	-0.023	07/02/2021	09:29

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-070221 WorkList ID : 150292 Department : Wet-Chemistry Date : 07-02-2021 08:21:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/02/2021	Water	M2920-01	Turbidity	Cool 4 deg C	LOCK01	J51	MW-2	06/30/2021	SM2130 B
07/02/2021	Water	M2920-03	Turbidity	Cool 4 deg C	LOCK01	J51	MW-4	06/30/2021	SM2130 B

Date/Time 7/2/21 8:30
Raw Sample Received by: AD (WCC)
Raw Sample Relinquished by: JD (CSM)

Date/Time 7/2/21 10:15
Raw Sample Received by: JD (CSM)
Raw Sample Relinquished by: AD (WCC)