


LB115197

7471B

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.0711e-004

C= 2.4844e-002

Rho= 0.9999166

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.030  | 0.030  | 25    | 0.000      | 25    |       |       |       |       | -  |
| 0.2    | 0.200  | 0.229  | 0.029  | 987   | 0.0 %      | 987   |       |       |       |       | 15 |
| 2.5    | 2.500  | 2.452  | -0.048 | 11718 | 0.0 %      | 11718 |       |       |       |       | -2 |
| 5.0    | 5.000  | 4.993  | -0.007 | 23989 | 0.0 %      | 23989 |       |       |       |       | 0  |
| 7.5    | 7.500  | 7.429  | -0.071 | 35750 | 0.0 %      | 35750 |       |       |       |       | -1 |
| 10.0   | 10.000 | 10.067 | 0.067  | 48486 | 0.0 %      | 48486 |       |       |       |       | +1 |

## LB115197 INSTRUMENT ID : CV1

| Method: 7471B    |                           | Operator: Admin |         | Date of Analysis: 02 Jul 2021 14:18:44 |        |       |                      |        |
|------------------|---------------------------|-----------------|---------|----------------------------------------|--------|-------|----------------------|--------|
| Sample ID        | Extended ID               | µ Abs.          | Conc.   | Std Cond                               | Method | Units | Date                 | Type   |
| 0.0 - 1          | 90                        | 25              | -       | 0.00007471B                            | PPB    |       | 02 Jul 2021 14:21:02 | S Std  |
| 0.2 - 1          | 50.2                      | 987             | -       | 0.20007471B                            | PPB    |       | 02 Jul 2021 14:23:18 | S Std  |
| 2.5 - 1          | 52.5                      | 11718           | -       | 2.50007471B                            | PPB    |       | 02 Jul 2021 14:25:36 | S Std  |
| 5.0 - 1          | 55                        | 23989           | -       | 5.00007471B                            | PPB    |       | 02 Jul 2021 14:27:52 | S Std  |
| 7.5 - 1          | 57.5                      | 35750           | -       | 7.50007471B                            | PPB    |       | 02 Jul 2021 14:30:09 | S Std  |
| 10.0 - 1         | 57.5                      | 48486           | -       | 10.00007471B                           | PPB    |       | 02 Jul 2021 14:32:25 | S Std  |
| ICV88 - 1        | ICV88                     | 18639           | 3.8852  | -7471B                                 | PPB    |       | 02 Jul 2021 14:35:06 | U SMPL |
| ICB88 - 1        | ICB88                     | -120            | -0.0000 | -7471B                                 | PPB    |       | 02 Jul 2021 14:37:21 | U SMPL |
| CCV11 - 1        | CCV11                     | 23669           | 4.9269  | -7471B                                 | PPB    |       | 02 Jul 2021 14:39:36 | U SMPL |
| CCB11 - 1        | CCB11                     | -113            | 0.0014  | -7471B                                 | PPB    |       | 02 Jul 2021 14:41:51 | U SMPL |
| CRA - 1          | CRA                       | 946             | 0.2208  | -7471B                                 | PPB    |       | 02 Jul 2021 14:44:07 | U SMPL |
| HighStd - 1      | HighStd                   | 48318           | 10.0320 | -7471B                                 | PPB    |       | 02 Jul 2021 14:46:22 | U SMPL |
| ChkStd - 1       | ChkStd                    | 33313           | 6.9243  | -7471B                                 | PPB    |       | 02 Jul 2021 14:48:37 | U SMPL |
| PB137494BL - 1   | PBS                       | -166            | -0.0095 | -7471B                                 | PPB    |       | 02 Jul 2021 14:50:55 | U SMPL |
| PB137494BS - 1   | LCSS                      | 19434           | 4.0498  | -7471B                                 | PPB    |       | 02 Jul 2021 14:53:12 | U SMPL |
| M2904-01 - 1     | COMP-1                    | 13049           | 2.7274  | -7471B                                 | PPB    |       | 02 Jul 2021 15:01:42 | U SMPL |
| M2904-01DUP - 1  | COMP-1DUP                 | 12559           | 2.6259  | -7471B                                 | PPB    |       | 02 Jul 2021 15:03:58 | U SMPL |
| M2904-01MS - 1   | COMP-1MS                  | 31933           | 6.6385  | -7471B                                 | PPB    |       | 02 Jul 2021 15:06:15 | U SMPL |
| M2904-01MSD - 1  | COMP-1MSD                 | 29824           | 6.2017  | -7471B                                 | PPB    |       | 02 Jul 2021 15:08:32 | U SMPL |
| M2904-03 - 1     | COMP-2                    | 4174            | 0.8893  | -7471B                                 | PPB    |       | 02 Jul 2021 15:10:48 | U SMPL |
| CCV12 - 1        | CCV12                     | 23498           | 4.8915  | -7471B                                 | PPB    |       | 02 Jul 2021 15:15:44 | U SMPL |
| CCB12 - 1        | CCB12                     | -138            | -0.0037 | -7471B                                 | PPB    |       | 02 Jul 2021 15:17:58 | U SMPL |
| M2904-04 - 1     | COMP-3                    | 2156            | 0.4714  | -7471B                                 | PPB    |       | 02 Jul 2021 15:20:14 | U SMPL |
| M2913-01 - 1     | D-2500                    | 7318            | 1.5405  | -7471B                                 | PPB    |       | 02 Jul 2021 15:22:30 | U SMPL |
| M2913-03 - 1     | D-2700                    | 10778           | 2.2571  | -7471B                                 | PPB    |       | 02 Jul 2021 15:24:45 | U SMPL |
| M2913-05 - 1     | D-3100                    | 8148            | 1.7124  | -7471B                                 | PPB    |       | 02 Jul 2021 15:27:00 | U SMPL |
| M2913-07 - 1     | D-3400                    | 3630            | 0.7767  | -7471B                                 | PPB    |       | 02 Jul 2021 15:29:16 | U SMPL |
| M2913-09 - 1     | D-3700                    | 8229            | 1.7291  | -7471B                                 | PPB    |       | 02 Jul 2021 15:31:32 | U SMPL |
| M2913-11 - 1     | D-4050                    | 16480           | 3.4380  | -7471B                                 | PPB    |       | 02 Jul 2021 15:33:49 | U SMPL |
| M2913-13 - 1     | D-4300                    | 14843           | 3.0990  | -7471B                                 | PPB    |       | 02 Jul 2021 15:36:07 | U SMPL |
| M2913-15 - 1     | P-D20                     | 63550           | 13.1867 | -7471B                                 | PPB    |       | 02 Jul 2021 15:38:24 | U SMPL |
| M2913-17 - 1     | P-D19                     | 63871           | 13.2531 | -7471B                                 | PPB    |       | 02 Jul 2021 15:40:40 | U SMPL |
| CCV13 - 1        | CCV13                     | 23955           | 4.9862  | -7471B                                 | PPB    |       | 02 Jul 2021 15:42:57 | U SMPL |
| CCB13 - 1        | CCB13                     | -127            | -0.0015 | -7471B                                 | PPB    |       | 02 Jul 2021 15:45:14 | U SMPL |
| M2917-01 - 1     | FK-BPM-04-017-ABCD        | 1174            | 0.2680  | -7471B                                 | PPB    |       | 02 Jul 2021 15:47:30 | U SMPL |
| M2918-01 - 1     | FK-BPM-03-023-ABCD        | 1733            | 0.3838  | -7471B                                 | PPB    |       | 02 Jul 2021 15:49:45 | U SMPL |
| M2919-01 - 1     | FK-BPM-01-B9-B9A-B10-B10A | 1097            | 0.2520  | -7471B                                 | PPB    |       | 02 Jul 2021 15:52:00 | U SMPL |
| M2904-01LX5 - 1  |                           | 2388            | 0.5194  | -7471B                                 | PPB    |       | 02 Jul 2021 15:54:16 | U SMPL |
| M2904-01A - 1    |                           | 32933           | 6.8456  | -7471B                                 | PPB    |       | 02 Jul 2021 15:56:32 | U SMPL |
| M2913-15DLX5 - 1 |                           | 12282           | 2.5686  | -7471B                                 | PPB    |       | 02 Jul 2021 15:58:49 | U SMPL |
| M2913-17DLX5 - 1 |                           | 12396           | 2.5922  | -7471B                                 | PPB    |       | 02 Jul 2021 16:01:06 | U SMPL |
| CCV14 - 1        | CCV14                     | 25918           | 5.3927  | -7471B                                 | PPB    |       | 02 Jul 2021 16:03:23 | U SMPL |
| CCB14 - 1        | CCB14                     | -114            | 0.0012  | -7471B                                 | PPB    |       | 02 Jul 2021 16:05:40 | U SMPL |