



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115199

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.01	07/02/2021	14:25
2	CAL2	1	Water	NA	NA	21.8	7.00	07/02/2021	14:26
3	CAL3	1	Water	NA	NA	21.7	10.03	07/02/2021	14:27
4	ICV	1	Water	NA	NA	21.6	7.01	07/02/2021	14:30
5	CCV1	1	Water	NA	NA	21.4	2.02	07/02/2021	14:31
6	M2920-01	1	Water	NA	NA	22.5	6.75	07/02/2021	14:32
7	M2920-01DUP	1	Water	NA	NA	22.5	6.76	07/02/2021	14:33
8	M2920-03	1	Water	NA	NA	22.7	6.76	07/02/2021	14:34
9	CCV2	1	Water	NA	NA	21.8	12.02	07/02/2021	14:35

WORKLIST(Hardcopy Internal Chain)

LB 115799

WorkList Name : PH7221

WorkList ID : 150293

Department : Wet-Chemistry

Date : 07-02-2021 08:53:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/30/2021	Water	M2920-01	pH	Cool 4 deg C	LOCK01	J51	MW-2	06/30/2021	SM 4500-H B
06/30/2021	Water	M2920-03	pH	Cool 4 deg C	LOCK01	J51	MW-4	06/30/2021	SM 4500-H B

Date/Time 07/02/21 14:15
Raw Sample Received by: KS (Wetchem)
Raw Sample Relinquished by: AS (Wetchem)

Date/Time 07/02/21 14:40
Raw Sample Received by: AS (Wetchem)
Raw Sample Relinquished by: KS (Wetchem)