


LB115201

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5917e-004

C= -2.8542e-002

Rho= 0.9998912

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.005	-0.005	90	0.000	90					-6
0.2	0.200	0.189	-0.011	838	0.0 %	838					1
2.5	2.500	2.525	0.025	9854	0.0 %	9854					-1
5.0	5.000	4.947	-0.053	19196	0.0 %	19196					1
7.5	7.500	7.603	0.103	29447	0.0 %	29447					-1
10.0	10.000	9.941	-0.059	38468	0.0 %	38468					-1

LB115201 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 02 Jul 2021 16:13:18					Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units			
0.0 - 1	50	90	-	0.0000	7470A	PPB	02 Jul 2021 16:13:43	S	Std
0.2 - 1	50.2	838	-	0.2000	7470A	PPB	02 Jul 2021 16:16:00	S	Std
2.5 - 1	52.5	9854	-	2.5000	7470A	PPB	02 Jul 2021 16:18:17	S	Std
5.0 - 1	55	19196	-	5.0000	7470A	PPB	02 Jul 2021 16:20:34	S	Std
7.5 - 1	57.5	29447	-	7.5000	7470A	PPB	02 Jul 2021 16:22:50	S	Std
10.0 - 1	510	38468	-	10.0000	7470A	PPB	02 Jul 2021 16:25:06	S	Std
ICV89 - 1	ICV89	15170	3.9031	-7470A	PPB		02 Jul 2021 16:27:49	U	SMPL
ICB89 - 1	ICB89	-62	-0.0446	-7470A	PPB		02 Jul 2021 16:30:03	U	SMPL
CCV15 - 1	CCV15	19665	5.0681	-7470A	PPB		02 Jul 2021 16:32:18	U	SMPL
CCB15 - 1	CCB15	-77	-0.0485	-7470A	PPB		02 Jul 2021 16:34:33	U	SMPL
CRA - 1	CRA	831	0.1868	-7470A	PPB		02 Jul 2021 16:36:48	U	SMPL
HighStd - 1	HighStd	38182	9.8672	-7470A	PPB		02 Jul 2021 16:39:03	U	SMPL
ChkStd - 1	ChkStd	25504	6.5814	-7470A	PPB		02 Jul 2021 16:41:18	U	SMPL
PB137493BL - 1	PBW	-110	-0.0571	-7470A	PPB		02 Jul 2021 16:43:34	U	SMPL
PB137493BS - 1	LCSW	15129	3.8925	-7470A	PPB		02 Jul 2021 16:45:50	U	SMPL
M2830-01 - 1	18B-CONCRETE	-33	-0.0371	-7470A	PPB		02 Jul 2021 16:48:06	U	SMPL
M2830-01DUP - 1	18B-CONCRETEDUP	48	-0.0161	-7470A	PPB		02 Jul 2021 16:50:22	U	SMPL
M2830-01MS - 1	18B-CONCRETEMS	13755	3.5364	-7470A	PPB		02 Jul 2021 16:52:39	U	SMPL
M2830-01MSD - 1	18B-CONCRETEMSD	14605	3.7567	-7470A	PPB		02 Jul 2021 16:54:56	U	SMPL
M2830-03 - 1	18B-ASPHALT	-69	-0.0464	-7470A	PPB		02 Jul 2021 16:57:10	U	SMPL
CCV16 - 1	CCV16	19643	5.0624	-7470A	PPB		02 Jul 2021 16:59:25	U	SMPL
CCB16 - 1	CCB16	-70	-0.0467	-7470A	PPB		02 Jul 2021 17:01:40	U	SMPL
M2913-02 - 1	D-2500	33	-0.0200	-7470A	PPB		02 Jul 2021 17:03:56	U	SMPL
M2913-04 - 1	D-2700	44	-0.0171	-7470A	PPB		02 Jul 2021 17:06:11	U	SMPL
M2913-06 - 1	D-3100	50	-0.0156	-7470A	PPB		02 Jul 2021 17:08:26	U	SMPL
M2913-08 - 1	D-3400	62	-0.0125	-7470A	PPB		02 Jul 2021 17:10:42	U	SMPL
M2913-10 - 1	D-3700	45	-0.0169	-7470A	PPB		02 Jul 2021 17:12:58	U	SMPL
M2913-12 - 1	D-4050	53	-0.0148	-7470A	PPB		02 Jul 2021 17:15:14	U	SMPL
M2913-14 - 1	D-4300	58	-0.0135	-7470A	PPB		02 Jul 2021 17:17:30	U	SMPL
M2913-16 - 1	P-D20	52	-0.0151	-7470A	PPB		02 Jul 2021 17:19:46	U	SMPL
M2913-18 - 1	P-D19	47	-0.0164	-7470A	PPB		02 Jul 2021 17:22:03	U	SMPL
CCV17 - 1	CCV17	19418	5.0041	-7470A	PPB		02 Jul 2021 17:24:18	U	SMPL
CCB17 - 1	CCB17	-61	-0.0444	-7470A	PPB		02 Jul 2021 17:26:33	U	SMPL
PB137493TB - 1	PB137481TB	42	-0.0177	-7470A	PPB		02 Jul 2021 17:28:48	U	SMPL
M2830-01LX5 - 1		69	-0.0107	-7470A	PPB		02 Jul 2021 17:31:04	U	SMPL
M2830-01A - 1		15488	3.9855	-7470A	PPB		02 Jul 2021 17:33:19	U	SMPL
CCV18 - 1	CCV18	19071	4.9142	-7470A	PPB		02 Jul 2021 17:35:35	U	SMPL
CCB18 - 1	CCB18	-91	-0.0521	-7470A	PPB		02 Jul 2021 17:37:51	U	SMPL