



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115202

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.01	07/02/2021	14:25
2	CAL2	1	Water	NA	NA	21.8	7.00	07/02/2021	14:26
3	CAL3	1	Water	NA	NA	21.7	10.03	07/02/2021	14:27
4	ICV	1	Water	NA	NA	21.6	7.01	07/02/2021	17:10
5	CCV1	1	Water	NA	NA	21.6	2.01	07/02/2021	17:11
6	M2928-01	1	Water	NA	NA	19.9	5.91	07/02/2021	17:12
7	M2928-01DUP	1	Water	NA	NA	19.9	5.92	07/02/2021	17:13
8	M2928-02	1	Water	NA	NA	20.2	5.87	07/02/2021	17:14
9	M2928-03	1	Water	NA	NA	19.8	6.03	07/02/2021	17:15
10	M2928-07	1	Water	NA	NA	20.6	8.81	07/02/2021	17:16
11	CCV2	1	Water	NA	NA	21.7	12.02	07/02/2021	17:17

WORKLIST(Hardcopy Internal Chain)

LB115202

WorkList Name : PH7221-1 WorkList ID : 150309 Department : Wet-Chemistry Date : 07-02-2021 17:01:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/02/2021	Water	M2928-01	pH	Cool 4 deg C	PSEG03	F51	BUR-1064	07/02/2021	9040C
07/02/2021	Water	M2928-02	pH	Cool 4 deg C	PSEG03	F51	BUR-1065	07/02/2021	9040C
07/02/2021	Water	M2928-03	pH	Cool 4 deg C	PSEG03	F51	BUR-1066	07/02/2021	9040C
07/02/2021	Water	M2928-07	pH	Cool 4 deg C	PSEG03	F51	DRUM-1067	07/02/2021	9040C

Date/Time 07/02/21 17:05
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: LB CHLJ

Date/Time 07/02/21 17:20
 Raw Sample Received by: KS (DLJ)
 Raw Sample Relinquished by: KS (WC)