



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/7/2021

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 07/06/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 07/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115206

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2927-01	AUD-1131	1	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2927-02	AUD-1132	2	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2927-03	AUD-1133	3	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2930-01	BUR-1061	4	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2930-02	BUR-1062	5	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2930-03	BUR-1063	6	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2934-01	WASTE	7	1.14	9.85	8.42	83.6	
M2934-02	VOC	8	1.16	9.64	8.28	84.0	
M2934-03	1	9	1.15	9.83	8.47	84.3	
M2934-04	2	10	1.11	9.77	8.11	80.8	
M2934-05	3	11	1.15	9.7	8.13	81.6	
M2934-06	4	12	1.13	9.76	8.37	83.9	
M2934-07	5	13	1.16	9.95	8.44	82.8	
M2935-01	WASTE	14	1.12	9.65	8.25	83.6	
M2935-02	VOC	15	1.13	9.94	8.34	81.8	
M2935-03	1	16	1.18	9.83	8.27	82.0	
M2935-04	2	17	1.16	9.86	8.39	83.1	
M2935-05	3	18	1.14	9.72	7.82	77.9	
M2935-06	4	19	1.11	9.95	8.71	86.0	
M2935-07	5	20	1.12	9.85	8.51	84.7	
M2935-07D	5DUP	21	1.13	9.87	8.52	84.6	
M2937-01	TRE-1220	22	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2940-01	LOD-MDL-SOIL-01-QT3-2021	23	1.00	2.00	2.00	100.0	
M2940-02	LOQ-SOIL-02-QT3-2021	24	1.00	2.00	2.00	100.0	
M2940-03	MDL-SOIL-03-QT3-2021	25	1.00	2.00	2.00	100.0	
M2940-04	LOD-MDL-SOIL-04-QT3-2021	26	1.00	2.00	2.00	100.0	
M2940-05	LOQ-SOIL-05-QT3-2021	27	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/7/2021

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 07/06/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 07/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115206

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2940-06	MDL-SOIL-06-QT3-2021	28	1.00	2.00	2.00	100.0	
M2940-16	LLOQ-SOIL-QT3-2021	29	1.00	2.00	2.00	100.0	
M2941-01	03-A-03-B-03-C	30	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M2942-01	41893	50	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2942-02	41894	51	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2942-03	41895	52	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2943-01	41896	53	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2943-02	41897	54	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2943-03	41898	55	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2944-01	BU-03-070621	56	1.11	9.97	9.43	93.9	
M2946-01	SB-1	31	1.14	9.9	9.12	91.1	
M2946-02	SB-2	32	1.14	9.83	9.39	94.9	
M2946-03	SB-3	33	1.12	9.96	9.28	92.3	
M2946-04	SB-4	34	1.11	9.78	9.05	91.6	
M2946-05	SB-5	35	1.15	9.84	9.18	92.4	
M2946-06	SB-6	36	1.11	9.85	9.3	93.7	
M2947-01	SOIL-BANK-1	37	1.14	9.71	8.83	89.7	
M2947-02	SOIL-BANK-2	38	1.18	9.64	8.69	88.8	
M2947-03	SOIL-BANK-3	39	1.15	9.88	9.02	90.1	
M2947-04	SOIL-BANK-4	40	1.11	9.52	8.61	89.2	
M2947-05	SOIL-BANK-5	41	1.12	9.8	8.69	87.2	
M2947-06	SOIL-BANK-6	42	1.17	9.66	8.93	91.4	
M2947-07	SOIL-PILE-7	43	1.11	9.8	9.02	91.0	
M2947-08	SOIL-PILE-8	44	1.14	9.95	9.14	90.8	
M2947-09	SOIL-PILE-9	45	1.14	9.61	9.05	93.4	
M2947-10	A4-SOIL-BANK	46	1.16	9.9	8.98	89.5	
M2947-11	B6-SOIL-BANK	47	1.19	9.8	9.06	91.4	
M2947-12	C7-SOIL-PILE	48	1.12	9.88	9.04	90.4	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/7/2021

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 07/06/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 07/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115206

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2947-13	D9-SOIL-PILE	49	1.13	9.7	9.14	93.5	
M2949-01	POLY-149TH-ST-STATION	57	1.00	2.00	2.00	100.0	PLASTIC SAMPLE
M2949-02	PPE-149TH-ST-STATION	58	1.00	2.00	2.00	100.0	PLASTIC SAMPLE
M2951-01	AR-FORT-1	59	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-02	AR-FORT-2	60	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-03	AR-FORT-3	61	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-04	AR-FORT-4	62	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-05	AR-FORT-5	63	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-06	AR-FORT-6	64	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-07	AR-FORT-7	65	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-08	AR-FORT-8	66	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-09	AR-FORT-9	67	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-10	AR-FORT-10	68	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-11	AR-FORT-11	69	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-12	AR-FORT-12	70	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-13	AR-FORT-13	71	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-14	AR-FORT-14	72	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-15	AR-FORT-15	73	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-16	AR-FORT-16	74	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-17	AR-FORT-17	75	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-18	AR-FORT-18	76	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-19	AR-FORT-19	77	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2951-20	AR-FORT-20	78	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-01	AR-FORT-21	79	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-02	AR-FORT-22	80	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-03	AR-FORT-23	81	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-04	AR-FORT-24	82	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-05	AR-FORT-25	83	1.00	2.00	2.00	100.0	WIPE SAMPLE



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/7/2021

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 07/06/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 07/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115206

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2952-06	AR-FORT-26	84	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-07	AR-FORT-27	85	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-08	AR-FORT-28	86	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-09	AR-FORT-29	87	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-10	AR-FORT-30	88	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-11	AR-FORT-31	89	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-12	AR-FORT-32	90	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-13	AR-FORT-33	91	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-14	AR-FORT-34	92	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-15	AR-FORT-35	93	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-16	AR-FORT-36	94	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-17	AR-FORT-37	95	1.00	2.00	2.00	100.0	WIPE SAMPLE
M2952-18	AR-FORT-38	96	1.00	2.00	2.00	100.0	WIPE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$